

Declaration of Conformity-V1.0

# Declaration of Conformity



**Manufacturer:** Shenzhen Mindray Scientific Co., Ltd.  
6/F, Bldg 2, 1203 Nanhuan Avenue, Yutang Block, Guangming District,  
518106 Shenzhen, P. R. China

**EC-Representative:** Shanghai International Holding Corp. GmbH (Europe)  
Eiffestraße 80, 20537 Hamburg, Germany

**Product:** Syringe pump (Including Accessories)

**Model:** BeneFusion eSP, BeneFusion eSP ex, BeneFusion eSP Neo

**Classification:** IIb (According to Rule 11 of MDD Annex IX)

**GMDN Code:** 13217

**MD Code:** MD 1101

**Conformity**  
MDD Annex II excluding (4)

**Assessment Route:**

We herewith declare under our sole responsibility that the above mentioned products meet the provisions of the Council Directive 93/42/EEC for Medical Device, as amended by 2007/47/EC. All supporting documentations are retained under the premises of the manufacturer.

## Standards Applied:

List of (harmonized) standards for which documented evidence for compliance can be provided as attachment.

**Notified Body:** TÜV SÜD Product Service GmbH  
Ridlerstraße 65  
80339 München, Germany

**Notified Body No. :** 0123

**Start of CE-Marking:** 2020.11.20

**Place, Date of Issue:** Shenzhen,

**Signature:**

..... *Bai Yanhong* 2020.11.20

**Name of Authorized Signatory:** Bai Yanhong

**Position Held in Company:** Manager, Technical Regulation

Attachment of Declaration of Conformity: Applied Standards List-V1.0

**Product:** Syringe pump

**Model:** BeneFusion eSP, BeneFusion eSP ex, BeneFusion eSP Neo

**Applied Standards:**

|                           |   |
|---------------------------|---|
| EN 60601-2-24:2015        | Medical electrical equipment -- Part 2-24: Particular requirements for the safety of infusion pumps and controllers   |
| EN 60601-1:2006/A1:2013   | Medical electrical equipment - Part 1: General requirements for basic safety and essential performance  |
| EN 60601-1-8:2007/A1:2013 | Medical electrical equipment - Part 1-8: General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems |
| EN 62366-1:2015           | Medical devices - Application of usability engineering to medical devices   |
| EN 60601-1-6:2010/A1:2015 | Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral Standard: Usability   |
| EN ISO 14971:2012         | Medical devices - Application of risk management to medical devices   |
| EN 62304:2006/A1:2015     | Medical device software - Software life-cycle processes   |
| EN 60601-1-2:2015         | Medical Electrical Equipment - Part 1-2 General Requirements for Safety - Collateral Standard: Electromagnetic compatibility-Requirements and tests   |
| EN 1041:2008+A1:2013      | Information supplied by the manufacturer of medical devices   |
| EN ISO 15223-1: 2016      | Medical devices -<br>Symbols to be used with medical device labels,<br>labeling, and information to be supplied -<br>Part 1: General requirements   |
| EN 1789:2007+A2:2014      | Medical vehicles and their equipment - Road ambulances  |