


PIONEER CLW

4 MOTORLU TARTILI, LİFTLİ YOĞUN BAKIM HASTA KARYOLASI
4 MOTORS COLUMN INTENSIVE CARE PATIENT BED WITH WEIGHT SCALE



Yükseklik / High Low



Diz Kırma / Knee Break



Ayak Hareketi / Legrest



Sirt - Ayak Hareketi / Auto Contour

PIONEER CLW

4 MOTORLU TARTILI, LİFTLİ YOĞUN BAKIM HASTA KARYOLASI
MOTORS COLUMN INTENSIVE CARE PATIENT BED WITH WEIGHT SCALE



TEKNİK ÖZELLİKLER

- Gelişmiş entegre tartı sistemi
- Elektronik motor yardımı ile yükseklik, başucu ve ayakucu hareketi.
- Trendelenburg, ters trendelenburg, fowler ve vasküler pozisyonları.
- Hasta el kumandası.
- Hemşire kontrol ünitesi.
- Elektronik CPR tuşu.
- Otomatik kayma sistemine sahip başucu
- Auto contour.
- Hijyenik dizayna sahip plastik (PP), kilitlenebilen ve gizlenebilen yan korkuluklar.
- Kolay sökülüp takılabilen ABS plastik muhafazalı, profil şase.
- Hijyenik dizayna sahip plastik (PP) kolay sökülüp takılabilen başlıklar.
- Antistatik tekerler.(merkezi ve doğrusal kilit)
- Elektrostatik boyalı metal aksam.
- Yüksekliği ayarlanabilir paslanmaz çelik, serum askısı.
- Diz hareketlerini kolay ayarlayabilme özelliği.
- Plastik çarpma tamponları.
- 24V akü ile kesintisiz çalışabilme.
- Sirt kısmında mekanik CPR.
- Başucu ve ayakucu köşelerinde aksesuar yuvası.
- Sirt açış göstergesi.
- Karyola açış göstergesi.
- Tek tuşla cardiac chair (Sandalye) pozisyonu.
- Tek tuşla shock pozisyonu.
- Tek tuşla semi fowler pozisyonu.
- Tek tuşla yataktan çıkış pozisyonu.
- Tek tuşla muayene pozisyonu.
- İstem dışı kullanımları önlemek için hemşire kumandasında aktivasyon tuşu.
- Yatak platform uzatması 18 cm
- Kompak laminant sırt ve X-Ray kaset yuvası.
- IPX4 elektronik sistem.

TECHNICAL DETAILS

- Comprehensive integrated weighing system.
- Electrical operated backrest, height and legrest adjustment.
- Electrical operated trendelenburg, reverse trendelenburg, fowler and vascular position.
- Patient handset.
- Nurse control unit.
- Auto-CPR.
- Retracting backrest for advanced comfort, less pressure.
- Auto contour.
- Hygienic design (PP), lockable and tuck away side rails.
- Profile frame with easy removable ABS plastic covers.
- Removable head and foot ends.
- Central and directional lockable castors.
- Electrostatic painted metal frame.
- Height adjustable stainless steel IV pole.
- Easy adjustable knee break position.
- Plastic crash bumpers.
- Battery Backup.
- Dual sided manual CPR at backrest.
- Comprehensive range of accessories socket at corners.
- Angle gauge for back.
- Angle gauge for bed.
- Cardiac chair position with one button.
- Shock position with one button.
- Semi-Fowler position with one button.
- Bed exit position with one button.
- Examination position with one button.
- Activation key at remote control to prevent any undesired positions.
- Bed extension 18 cm
- X-Ray translucent backrest and cassette carrier.
- IPX4 standard electronic system.

TARTI GÖSTERGESİ
SCALE INDICATORHEMŞİRE KONTROL ÜNİTESİ
NURSE CONTROL UNITHASTA EL KUMANDASI
PATIENT HANDSET

EBAT / DIMENSIONS

• Uzunluk / Overall Length Overall	: 216 cm
• En / Overall Width	: 102 cm
• Yükseklik(Şiltesiz) / Height Range	: 46 -82 cm
• Trendelenburg / Trendelenburg	: 0°-16°
• Ters Trendelenburg / Reverse Trendelenburg	: 0°-16°
• Sirt Kısmı(Maximum) / Backrest Angle (Max.)	: 70°
• Ayak Ucu Kısmı(Maximum) / Legrest Angle (Max.)	: 40°
• Teker Çapı / Castor Diameter	: 15 cm
• Yatak Alt Boşluğu / Under Bed Clearance	: 15 cm
• Güvenli Çalışma Kapasitesi / Safe Working Load	: 230 kg
• (Aksesuarlar dahil) / (With Accessories)	

ELEKTRONİK VERİLER / ELECTRICAL DATA

• Güç Girişi / Power In	: 1.5 A max. 230V AC 50-60 Hz
• Elektrik Çarpmasına Karşı Koruma / Electrical Shock Protection	: Class I, Type B
• Su Sızmasına Karşı Koruma / Liquid Ingress Protection	: IPX4 - IPX6
• Akü / Batteries	: 2x 12V Series, Connected, Sealed, Rechargeable, Lead / Acid gel, 1.2 amp hrs.

MERKEZİ FREN SİSTEMİ
CENTRAL LOCKSirt Hareketi
BackrestAyak Hareketi
LegrestSirt-Ayak Hareketi
Auto ContourDiz Kırma
Knee BreakYükseklik
High/LowDüz Trendelenburg
TrendelenburgTers Trendelenburg
Reverse TrendelenburgKardiyak
Cardiac Chair



CERTIFICATE

CERTIFICATE Nr. SQ/03-2020

WE HEREBY CERTIFY THAT THE QUALITY MANAGEMENT SYSTEM OPERATED BY

MEYSA TIBBİ CİHAZLAR SANAYİ VE TİCARET A.Ş

Registered office address:
O.S.B 16. Cad. No:73 MELİKGAZİ KAYSERİ (TURKEY)

Certified operative unit address:
O.S.B 16. Cad. No:73 MELİKGAZİ KAYSERİ (TURKEY)

**IS IN COMPLIANCE WITH THE REQUIREMENTS OF STANDARD
UNI CEI EN ISO 13485:2016**

THIS CERTIFICATE IS VALID FOR THE FOLLOWING ACTIVITIES
Design and production of hospital furnitures, medical bed and patiend bed

FIRST ISSUE 31/05/2019 CURRENT ISSUE 01/06/2020 REVISION Nr. 02 EXPIRING DATE 01/06/2021



SGQ N° 047 A

Membro degli Accordi di Mutuo Riconoscimento
EA, IAF e ILAC

Signatory of EA, IAF and ILAC
Mutual Recognition Agreements

MTIC INTERCERT S.r.l. - Via Moscova, 11 - 20017 RHO (MI) - ITALY
www.mticert.org

Dipl.- Ing. Feridoon Sergizzarea
MTIC INTERCERT Certification Body



MEYSA TIBBİ CİHAZLAR SAN. ve TİC. A.Ş.

O. S. B. 16.Cad. No :73 Melikgazi / KAYSERİ / TÜRKİYE

+90 352 503 16 60

www.meissa.com.tr

+90 352 503 16 61

info@meissa.com.tr



DECLARATION OF CONFORMITY

IN accordance with 93/42/EC Medical Devices Directive of the Council of European Union, whose purpose is providing conformity of laws, directives and administrative of documents of member countries in respect to Medical Device;

Name and Address of the Company: MEYSA TIBBİ CİHAZLAR SAN VE TİC A.S.

Tel: +90 352 5013 16 6 **Fax:** +90 352 503 16 61

Place and Address of Manufacture: OSB. 16.Cad No:73 Melikgazi Kayseri / TURKEY

Applied Directives: 93/42/EC- Medical Devices Directive

Classification and Annex Applied: Product is subject to Medical Devices Directive Class 1. Applied.

Name of Product and Types:

Patient and ICU Beds: PIONEER CLW, CL, 4, 4.3PLUS, 4.2CARE, 4.1TREND, 3, 2, ELITE CL, 4.3 PLUS,4, 3, 2 GRANDE CL, 4 ,3, 2, DUO

Manual Patient Beds: PRIMITIVE 3,2

Stretchers: MOVEE2.1, MOVEE 2.0, DYNAMIC 2.0, MOVEE VERSUS

Gynecological Ex. Table: Mathercare 2

Pediatric Bed: JUNIOR 4, 3, 2, JUNIOR, JUNIOR ELITE

Baby Cot: NICE 1, NICE 2

Examination Table: Diagno 01,02,03

Patient Bed Side Cabinets: SERA 1,2,3,4,5,6,7,8

Patient Overbed Table: MIRACH 1,2,3

Declaration;

Our company manufactures the products stated above in accordance with the requirements of the current EN 980-1996/A1 (Graphs and Symbols Used on labels) EN 1401 (Information provided with the Product by Manufacturer) ISO 13485:2003 Medical Devices Quality Management System.

Used Standards;

The mentioned products are complying with the requirements of the following **standards;** EN 60601-1:2006/A1:2013 EN 60601-1-2:2007 EN 60601-1-6:2010, IEC60601-2-38, EN 60601-2-52: 1-2010

The products described above were subjected to initial type experiments by Manufacturer and factory manufacture control was carried out by regular tests.

Date of Issue 01.06.2018

General Manager

