

EC Certificate

Directive 98/79/EC Annex IV, excluding Sections 4 and 6
Full Quality Assurance System
In Vitro Diagnostic Medical Devices

Registration No.: HL 60119814 0001

Report No.: 21265422 001

Manufacturer: Macherey-Nagel GmbH & Co. KG
Neumann-Neander-Str. 6-8
52355 Düren
Deutschland

Products: Products for self-testing
(see attachment for products and sites included)
Replaces Certificate, Registration No.: HL 60076687 0001

Expiry Date: 2022-05-28

The Notified Body hereby declares that the requirements of Annex IV, excluding section 4 and 6 of the directive 98/79/EC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex IV, section 5 of the aforementioned directive. For placing on the market of List A devices covered by this certificate an EC design-examination certificate according to Annex IV, section 4 and a verification of manufactured products according to section 6 is required.

Effective Date: 2017-05-29

Date: 2017-05-29

Notified Body


Dipl.-Ing. Sven Hoffmann



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 98/79/EC concerning in vitro diagnostic medical devices with the identification number 0197.

TÜV Rheinland
LGA Products GmbH
Tillystraße 2, 90431 Nürnberg

**Attachment to
Certificate**

Registration No.: HL 60119814 0001
Report No.: 21265422 001

Manufacturer: Macheray-Nagel GmbH & Co. KG
Neumann-Neander-Str. 6-8
52355 Düren
Deutschland

Products for self-testing:

- Single and multi-parameter disposable test strips for urine analysis
- indicator test strips and papers for measurement of pH in urine

Additional site for warehousing and logistics:

Bahnstr. 120
52355 Düren, Germany

Date: 2017-05-29

Notified Body


Dipl.-Ing. Sven Hoffmann

