



Bio-Rad  
Laboratories

Diagnostics Group  
9500 Jeronimo Road  
Irvine, California 92618-2017  
Telephone: (949) 598-1200

### CERTIFICATE-OF-ANALYSIS

Product Name: EQAS Serum Proteins Program  
 Master Batch Number: 183400  
 Product Base: Human Serum  
 Physical Form: Lyophilized  
 Date of Manufacture: 2022-11-30

Description	Material Number	Batch Number	Expiration Date
Cycle 50	BC23	183400	2024-02-29

Testing Results:

Each human donor unit used to manufacture this product was tested and found non-reactive at the donor level per current applicable FDA requirements using FDA-accepted methods including testing for HIV-1/HIV-2 antibody, HBsAg, HCV Antibody, HIV-1 and HCV by a nucleic acid test (NAT).

The bovine source material(s) of this product is collected in USDA licensed establishments. These animals received ante and post mortem inspections at the abattoir by a US veterinary service inspector.

Water Activity:	Specification: < 0.200aw	Level 1 < 0.030 aw	Level 2 < 0.030 aw	Level 3 < 0.030 aw	Level 4 < 0.030 aw
-----------------	--------------------------	-----------------------	-----------------------	-----------------------	-----------------------

Homogeneity Claim:

"Testing has been conducted to verify sufficient homogeneity in accordance with established requirements."

ADDITIONAL TESTING:
---------------------


This product has been manufactured under applicable guidelines/standards and meets all established Bio-Rad quality requirements.


Approved By:

Date:

  
 \_\_\_\_\_  
 (QC Supervisor/Manager)

  
 \_\_\_\_\_  
 12/14/22

  
 \_\_\_\_\_  
 (QA Supervisor/Manager)

  
 \_\_\_\_\_  
 12/15/2022

