

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: Urinalysis Reagent Strips

PURCHASE NO.: 37740068B

BATCH NO.: URS0080059

EXPIRATION DATE: 2022-09-19

QUANTITY: 180,000pcs

ANALYSIS		SPECIFICATIONS (Color card)	TEST RESULT
Ascorbic acid	Neg. urine	Negative	Complies
	20mg/dl	>10mg/dl <40mg/dl	Complies
Glucose	Neg. urine	Negative	Complies
	250mg/dL	>100mg/dL <500mg/dL	Complies
Bilirubin	Neg. urine	Negative	Complies
	Yihua Urine controls	>++	Complies
Ketone	Neg. urine	Negative	Complies
	Yihua Urine controls	>40mg/dL <160mg/dL	Complies
S.G	1.000	$\geq 1.000 < 1.005$	Complies
	1.025	$> 1.020 < 1.030$	Complies
Blood	Neg. urine	Negative	Complies
	Yihua Urine controls	$\geq ++$ $\leq +++$	Complies
pH	5.0	$\geq 5.0 < 6.0$	Complies
	9.0	$> 8.0 \leq 9.0$	Complies
Protein	Neg. urine	Negative	Complies
	Yihua Urine controls	$\neq +, \leq +++$	Complies
Urobilinogen	Neg. urine	Negative	Complies
	4~12mg/dL	$> 4\text{mg/dl}$ $\leq 12\text{mg/dl}$	Complies
Nitrite	Neg. urine	Negative	Complies
	Yihua Urine controls	+	Complies
Leukocyte	Neg. urine	Negative	Complies
	Yihua Urine controls	$> +$ $< +++$	Complies

Quality Department: Hedy Zhang

Date: Sep. 23, 2020