


Test Report

– Eppendorf PCR clean –

PCR Tube 0.2 mL



1000 Tubes 0.2 mL, thin walled
PCR clean
US-PAT.NO. 5,863,791

LOT J191924M
2025-07-28

Order no.: 0030 124.332
Eppendorf AG
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 951010006
Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

• Human DNA

Method: real-time PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• DNase

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable (LOD: 1.0×10^{-6} Kunitz units)

• PCR inhibition

Method: real-time PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable

• RNase

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable (LOD: 1.0×10^{-9} Kunitz units)


Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 09.09.20

Testing of the above described lot showed conformity within the limits of detection.

15.09.20
Date of release

Epp2009 7348
Registration number


Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.