



CHANGZHOU SHUANGMA MEDICAL DEVICES CO., LTD


Declaration

Hereby, we **Changzhou Shuangma Medical Devices Co., Ltd**, based in **San He Kou Development Zone, Zhenglu, Tianning, Changzhou** confirm technical description of our products:

| Lot nr. | Product name | Description | Picture |
|---------|--|--|---------|
| 1 | Sterile syringe, needle 25G x 1", 1 ml | <p>Sterile, disposable syringe</p> <ul style="list-style-type: none">- needle 25G x 1"- volume 1ml, degrees of volume measurement- with gradation from 0.1 ml to 0.1 ml- consisting of 3 parts: transparent cylinder, piston, rubber gasket- two-stage piston seal, which does not allow the solution to flow back during administration- made tightly- hypoallergenic needle, non-pyrogenic- individual packs <p>The reduced dead space of the syringe combined with the low dead space of the needle is 0,0044 ml (4,4 mkl).</p> <p>Packaged and labeled in accordance with Directive 93/42 / EEC (modified by Directive 2007/47 / EC), with indicating the name and address of the manufacturer, the name and the product code, the batch number, the batch production date, the batch expiry date, the country of origin, the sterilization method. Corresponding icons accompanied by relevant performance levels and reference to EN standard.</p> | |



CHANGZHOU SHUANGMA MEDICAL DEVICES CO., LTD

| | | | |
|---|--|---|--|
| 2 | Sterile syringe, needle 25G x 1" 1 ml | <p>Sterile, disposable syringe</p> <ul style="list-style-type: none">- needle 25G x 1"- volume 1ml, degrees of volume measurement- with gradation from 0.1 ml to 0.1 ml- consisting of 3 parts: transparent cylinder, piston, rubber gasket- two-stage piston seal, which does not allow the solution to flow back during administration- made tightly- hypoallergenic needle, non-pyrogenic- individual packs <p>Packaged and labeled in accordance with Directive 93/42 / EEC (modified by Directive 2007/47 / EC), with indicating the name and address of the manufacturer, the name and the product code, the batch number, the batch production date, the batch expiry date, the country of origin, the sterilization method. Corresponding icons accompanied by relevant performance levels and reference to EN standard.</p> | <p>样品照片 Sample Photo</p>  |
|---|--|---|--|

Place: Changzhou, China

Date: 2021/9/14

Signed: Lio

