

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: H.pylori IgG EIA Test Kit

PURCHASE NO.: 37740068A

BATCH NO.: 2003703

EXPIRATION DATE: 2021/10/14

QUANTITY: 15kits

SPECIFICATION		STANDARD	CONCLUSION
Physics Check		No sediment or floccule	Complies
Blank Well		OD < 0.050 (450nm/630nm) or OD < 0.100 (450nm)	Complies
Calibrators	Calibrator 1#	OD < 0.100	Complies
	Calibrator 2#	OD > 0.150	Complies
	Calibrator 6#	OD > 1.000	Complies
	R	≥ 0.98	Complies
Sensitivity	Negative	Conc. < 5AU/mL(8/8)	Complies
	Positive	Conc. > 5AU/mL(5/5)	Complies
	Minimum Detection Limit	Conc. < 5AU/mL	Complies
PRECISION		CV < 15%	Complies

Quality Department: Hedy Zhang

Date: Sep. 23, 2020

