



VERIFICATION OF CONFORMITY

S/N: 011099

No.: **ICR/VC/HM2410171**

Name and address of Applicant: Hengshui Deguang Medical Equipment Co., Ltd
No. 4, Row 4, No. 108, Wei Qi Road, Weitun Town, Binhu New District, Hengshui City, Hebei Province, China

Name and address of manufacturer: Hengshui Deguang Medical Equipment Co., Ltd
No. 4, Row 4, No. 108, Wei Qi Road, Weitun Town, Binhu New District, Hengshui City, Hebei Province, China

Product name: Hospital beds

Product types: QQ-S-I, QQ-S-II, QQ-Y-I, QQ-Y-II, ZQ-S-II, ZQ-Y-I, ZQ-Y-I, ZQ-Y-II, PC-I, PC-II, PC-III, A01-I, A01-II, A01-III, A01-IV, A02-I, A02-II, A02-III, A02-IV, B01-I, B01-II, B01-III, B01-IV, B02-I, B02-II, B02-III, B02-IV, B03-I, B03-II, B03-III, B03-IV, B01-III, ZQ-S-I, DG-XYC, DG-D01, DG-D02, DG-D03, DG-D04, DG-S01, DG-S02, DG-S03, DG-S04, DG-S05, DG-S06, DG-S07, DG-S08, DG-A01, DG-A02, DG-A03, DG-A04

Product trademark: DG

Verification was carried within following scope:

Information on the Declaration of Conformity:

Result:	Legislation:	Standard:
✓	MDR [(EU) 2017/745]	EN 60601-1:2006+A2:2021 EN 60601-1-2:2015+A1:2021 EN 60601-2-52:2010+A1:2015 EN 14971:2019

The assessment process has been carried out in accordance with individual rules and conditions agreed with the applicant. Evaluation has been carried out in accordance with:

Test report: UNICO2410DG70
Tests conducted by: Unico Certification Testing Co., Ltd
Issue date: 29.10.2024
Expiration date: 28.10.2029

Remarks:

- VoC was issued on voluntary basis and does not imply meeting all essential requirements listed in Declaration of Conformity.
- For introducing this product on European market may be needed EC/EU-type examination conducted by appropriate Notified Body.

ICR Co. Ltd.
www.icrqa.com
www.icrpolska.com
cert@icrqa.com



CEO, ICR Co., Ltd.

