

CERTIFICATE OF ANALYSIS

F-548S Phusion Flash High-Fidelity PC R Master Mix, 100 react.

Packaging Lot: 3504712

Expiry Date: 31.03.2027 (DD.MM.YYYY)

Storage: at -20±5°C

Filling lots for components in package:

| Lot | Quantity | Description |
|---------|----------|-----------------------------|
| 3458016 | 1 mL | 2X Phusion Flash Master Mix |

QUALITY CONTROL

| Parameter | Method | Requirement | Result |
|--------------------|--|---|----------|
| Functional testing | PCR amplification of 7.5 kb fragments from human genomic DNA and 20 kb fragment from lambda DNA and analysis on agarose gel after ethidium bromide staining. | Reactions produce specific PCR products | Conforms |

ISO CERTIFICATION

Manufactured by Thermo Fisher Scientific Baltics UAB, in compliance with ISO 9001 and ISO 13485 certified quality management system.

Quality authorized by:

E. Laveckis

