

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: HBsAg EIA Test Kit

PURCHASE NO.: 37740068A

BATCH NO.: 2004721

EXPIRATION DATE: 2021/11/12

QUANTITY: 50kits

SPECIFICATION		STANDARD	CONCLUSION
Physics Check		No visible deposit or floccules in solution.	Complies
Blank well		OD < 0.100 (450nm) OD < 0.050 (450/630)	Complies
Sensitivity	Ad0.2 ng/ml	OD ≥ 0.100	Complies
	Ad0.5 ng/ml	OD ≥ 0.150	Complies
	Ad1.0 ng/ml	OD ≥ 0.250	Complies
	Ay0.2 ng/ml	OD ≥ 0.100	Complies
	Ay0.5 ng/ml	OD ≥ 0.150	Complies
Coincidence of Specificity Standard		100/100	Complies
High Positive Standard		OD ≥ 3.000	Complies
Negative Control		OD < 0.100	Complies
Positive Control		OD > 1.000	Complies
Precision		CV < 15%	Complies

Quality Department: Hedy Zhang

Date: Sep 23, 2020

