

## **QUALITY CONTROL**

## **TECHNICAL DATA SHEET**

Document No : Validity Date : Revision No : Revision Date :

: SÜÖ.KK.63 : 15.01.2015 : 00 : -

Revision Date : Revision Reason : Page No : 1/

ITEM : AUTOMATIC HOLDER

**DEFINITION**: Automatic holder is a connector for holding comfortably of blood collection needle for collecting human

venous blood specimen.

Standard

## **SPEICIFCATION:**

Physical Properties	(Method of Test)	<u>Unit</u>	Result
Material	AYSET	-	Shall be medical material.
Visual Inspection	AYSET	-	Shall not have sharp edges or surface roughness capable of accidentally cutting and shall not have breaks or cracks in the material
Harmony between parts	AYSET	-	Shall be compatible
Resistance	AYSET	_	Shall be no visible error
Cleanliness	AYSET	-	Shall be clean.

## **Transport and Packing Properties**

Type of Packing	AYSET	<ul> <li>Shall be microwave cardboard, offset printed box and expressions specified in the standard</li> </ul>	
Transport–Storage Container	AYSET	<ul> <li>Inner box is put into carton/outer boxes.</li> </ul>	
Storage Condition	AYSET	<ul> <li>Keep away from direct sun light and stored in a clean and dry storage areas at max. 45 °C.</li> </ul>	





DISTRIBUTION:		PREPARED BY	APPROVED BY
Responsible Man. Quality Control Dept. Marketing Dept.	NAME & SURNAME	Ömer DEĞİRMENCİOĞLU	Özgül UYAN
	TITLE	Quality Control Manager	Responsible Manager
	DATE	15.01.2015	15.01.2015
	SIGNATURE		