



Benannt durch/Designated by:
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zfz.de
BS-MDR-099



Product Service

EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 010578 0039 Rev. 09

Manufacturer:

Drägerwerk AG & Co. KGaA

Moisliger Allee 53-55
23542 Lübeck
GERMANY

SRN Manufacturer:

DE-MF-000005329

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s). The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result. The quality management system assessment was accompanied by the assessment of technical documentation for devices selected on a representative basis. The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. The surveillance assessment shall also include an assessment of the technical documentation for the device or devices concerned on the basis of further representative samples. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G10_010578_0039_Rev_09

Report No.: 713253108_CN

Preceding Certificate No.: G10 010578 0039 Rev. 08

Valid from: 2023-03-14

Valid until: 2025-03-17

Date of Initial Issuance: 2020-03-18

Issue date: 2023-03-14

C. Dicks

Christoph Dicks

Head of Certification/Notified
Body





Benannt durch Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zglm.de
BS-MDR-099



Product Service

EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 010578 0039 Rev. 09

Classification:	Class IIa
Device Group:	R02 - BREATHING CIRCUITS AND CATHETER MOUNTS R0301 - RESPIRATORY MASKS R030201 - VENTILATION BALLOONS R0401 - VENTILATION FILTERS R0402 - NATURAL BREATHING FILTERS Z120301 - ANAESTHESIA AND PULMONARY VENTILATION SUPPORT INSTRUMENTS Z120309 - MEDICAL/MEDICINAL GAS PIPELINE SYSTEMS AND RELATED ACCESSORIES
Intended Purpose:	-
Classification:	Class IIa
Device Group:	Z12040192 - GENERAL MEDICINE DIAGNOSIS AND MONITORING INSTRUMENTS - MEDICAL DEVICE SOFTWARE
Intended Purpose:	-
Classification:	Class IIa
Device Group:	Z120302 - VITAL SIGNS MONITORING INSTRUMENTS
Intended Purpose:	-
Classification:	Class IIa
Device Group:	A060304 - INTRA-OPERATION FLUID COLLECTION DEVICES
Intended Purpose:	-
Classification:	Class IIb
Device Group:	Z12040192 - GENERAL MEDICINE DIAGNOSIS AND MONITORING INSTRUMENTS - MEDICAL DEVICE SOFTWARE
Intended Purpose:	Software intended to provide clinical information for the purpose of supporting patient management and the decision making process
Classification:	Class IIb
Device Group:	Z120804 - NEONATOLOGY INSTRUMENTS





Benannt durch Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zl.de
BS-MDR-099



Product Service

EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 010578 0039 Rev. 09

Intended Purpose: Warming therapy devices intended to provide controlled ambient conditions for premature babies and neonates in closed and open care therapy

Classification: Class IIb

Device Group: Z120301 - ANAESTHESIA AND PULMONARY VENTILATION SUPPORT INSTRUMENTS

Intended Purpose: Devices for the purpose of ventilation and/or anesthesia

Classification: Class IIb

Device Group: Z120309 - MEDICAL/MEDICINAL GAS PIPELINE SYSTEMS AND RELATED ACCESSORIES

Intended Purpose: Devices intended to distribute or supply gases, vacuum, electricity or data to equipment in diagnostic, therapy or surgery

Classification: Class IIb

Device Group: R020107 - THERMOREGULATED BREATHING CIRCUITS

Intended Purpose: Inspiratory (and expiratory) heated disposable breathing circuit for conducting humidified breathing gas from humidifier to patient

Classification: Class IIb

Device Group: R020101 - STANDARD BREATHING CIRCUITS

Intended Purpose: Devices intended to administer gases for the purpose of ventilation

The validity of this certificate ./.
depends on conditions and/or
is limited to the following:

Revision History:

Rev.	Dated	Report	Description
00	2020-03-18	713169482	-
01	2021-07-02	713184148	-
02	2021-09-30	713215188	-
03	2021-10-01	713215832	-
04	2021-10-04	713215842	-
05	2021-10-04	713219421	-
06	2021-11-22	713229134	-

Page 3 of 4

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



TUV®



Benannt durch Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zfl.de
BS-MDR-099



Product Service

EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class IIa and Class IIb Devices)

No. G10 010578 0039 Rev. 09

07	2022-02-21	713213004
08	2022-10-06	713225304_CN
09	2023-03-14	713253108_CN

-

-

Supplemented: Device(s)/group of
device(s) added



Drägerwerk AG & Co. KGaA, 23542 Lübeck, Germany

To whom it may concern

Manufacturer's Authorization

May 25, 2023

We, Drägerwerk AG & Co. KGaA, Moislinger Allee 53-55, 23558 Lübeck, Germany, who is an established and reputable manufacturer of medical equipment, having factories at Lübeck (Germany), Telford (United States), Andover (United States) and Shanghai (China), do hereby declare that

"Echipamed-Plus" SRL Valea Trandafirilor 24 "B", of. 2-7 MD-2001, Chisinau Republic of Moldova

is our official distributor and local representative for Anesthesia Devices, Incubators, Ventilation Machines, Monitoring, Pendant and Operating Lights of Drägerwerk AG & Co. KGaA in the territory of the Republic of Moldova.

We declare that only above mentioned Company is authorized to do registration, quote, sell, subsequently negotiate and sign contracts, as well as to perform installation and after sales Service of Anesthesia Devices, Incubators, Ventilation Machines, Monitoring, Pendant and Operating Lights manufactured by us in their own name and on their own account.

This authorization letter will remain valid until 31.12.2024.

Martin Koch
Managing Director Sub Region East Europe



Drägerwerk AG & Co. KGaA
Moislinger Allee 53-55
23558 Lübeck, Germany
Postal address:
23542 Lübeck, Germany
Tel. +49 451 882-0
Fax +49 451 882-2080
info@draeger.com
www.draeger.com
UID-Nr. DE135082211

Bank details:
Commerzbank AG, Lübeck
IBAN: DE95 2304 0022 0014 6795 00
Swift-Code: COBA DE FF 230
Sparkasse zu Lübeck
IBAN: DE15 2305 0101 0001 0711 17
Swift-Code: NOLADE21SPL

Registered office: Lübeck
Commercial register:
Local court Lübeck HRB 7903 HL
General partner:
Drägerwerk Verwaltungs AG
Registered office: Lübeck
Commercial register:
Local court Lübeck HRB 7395 HL

Chairman of the Supervisory Board
for Drägerwerk AG & Co. KGaA
and Drägerwerk Verwaltungs AG:
Stefan Lauer
Executive Board:
Stefan Dräger (chairman)
Rainer Klug
Gert-Hartwig Lescow
Dr. Reiner Piske
Anton Schrofner