

## Certificate of Conformity

Lot Number: 60848248

Part Number: 3451

Item Description: HLT 3451 TUBE 1.5 ML MCTGRADUATED CLEAR STERILE

Manufacturing Date: 07/02/2024

Irradiation Date: 07/17/2024

Sterility Assurance Level (SAL):  $10^{-6}$

Sterility Assurance Level for sterilized products is ensured by irradiating the product according to ANSI/AAMI/ISO 11137- 2015 guidelines.

Molecular BioProducts Inc., a part of Thermo Fisher Scientific Inc., hereby certifies that these products are manufactured in a controlled environment in accordance with established manufacturing guidelines, quality requirements and product specifications. They are certified, by validation of the manufacturing processes, to be free of the following contaminants:

- Human and mouse DNA (Deoxyribonucleic Acid): < 30 pg
- Ribonuclease (RNase): <  $10^{-9}$  Kunitz Units/ $\mu$ l
- Deoxyribonuclease (DNase): <  $10^{-7}$  Kunitz Units/ $\mu$ l
- PCR Inhibitors: not present
- Endotoxin (Pyrogen) < 0.06 EU/ml
- Adenosine-5'-triphosphate (ATP) <  $10^{-13}$  mg/ $\mu$ l



Certified by: \_\_\_\_\_ 08/13/2024  
Alice Chen

Inspector Leader

Laboratory Products Division  
Document No.: FRM0033085 Rev  
J.00

9389 Waples Street  
Area: Quality

San Diego, CA  
92121-3903

(800) 995-2787  
(858) 453-4367 fax

[www.thermofisher.com](http://www.thermofisher.com)  
ISO 13485:2016