



## MIRIS Heater™

### Technical specifications

<b>Dimensions (HxWxL)</b>	200x215x335 mm
<b>Weight</b>	3 kg
<b>Power supply</b>	110 V/240 V, 50/60 Hz
<b>Ambient Temperature</b>	5 to 40°C
<b>Maximum relative humidity</b>	80% R.H. in room temperatures up to 31°C decreasing linearly to 50 % R.H. at 40°C
<b>Operating Environment</b>	Indoor use only
<b>Temperature range</b>	5°C above ambient to 80°C
<b>Setting scale</b>	10 to 80°C in 0.5°C steps
<b>Temperature stability</b>	± 0.5°C
<b>Standards</b>	EN 61010-2-010 UL 61010A-2-010 CAN/CSA-C22.2 NO. 61010-2-010-04

Samples should be warmed to 40°C prior to analysis with Miris HMA™. The entire sample should be 40°C and the Miris HMA™ will also operate at 40°C.

