



INDUS MEDICARE PRIVATE LIMITED

“SURYODAYA”, 1-10-60/3, BEGUMPET, HYDERABAD – 500 016, TELANGANA, INDIA.

TEL : +91 40 27762603

FAX : +91 40 27761418

Q.A. LOT RELEASE TEST REPORT

DATE : 27.08.2022 REPORT NO. : IMPL / 22 / 304
CUSTOMER : DITA ESTFARM LLC – MOLDOVA LOT NO. : 22 06 001
QUANTITY : 1131.33 GROSS MFD. : 2022 - 06
TYPE OF PRODUCT: NATURAL, MALE LATEX CONDOMS (THICK 53mm) EXP. : 2027 - 05
BRAND NAME : ANY TIME 12'S SPECIFICATION: EN ISO 4074 : 2015

SL. NO	CRITERIA / TEST	INSPECTION LEVEL	AQL	SPECIFICATIONS	SAMPLE QTY. IN PIECES	NO. OF PIECES FAILED	MEAN	MINIMUM	MAXIMUM
1.	DIMENSIONS LENGTH NOMINAL WIDTH THICKNESS	13 PIECES	ACCEPT - 0 REJECT - 1	L = min.190 mm W = 53 ± 2 mm T = Min 0.100 mm	13	0	196.8 52.7 0.113	195.0 52.5 0.111	199.0 53.0 0.119
2.	BURSTING VOLUME AND PRESSURE BEFORE OVEN CONDITIONING	G - 1	1.5 ACCEPT - 10 REJECT - 11	B.V = min.18.0 ltrs. B.P = min. 1.0 KPa.	315	1	45.9 3.45	17.0 1.45	60.0 4.05
3.	FREEDOM FROM HOLES	G - 1 (min. 315pieces)	0.25 ACCEPT - 2 REJECT - 3	AS PER AQL	315	0	COMPLIES		
4.	VISIBLE DEFECTS	G - 1 (min. 315pieces)	0.4 ACCEPT - 3 REJECT - 4	AS PER AQL	315	0	COMPLIES		
5.	VISIBLY OPEN SEALS	G - 1 (min. 315pieces)	0.4 ACCEPT - 3 REJECT - 4	AS PER AQL	315	0	COMPLIES		
6.	PACKAGE INTEGRITY	S - 3	2.5 ACCEPT - 2 REJECT - 3	AS PER AQL	32	0	COMPLIES		
7.	LUBRICANT PLUS POWDER	13 PIECES	ACCEPT - 0 REJECT - 1	Min.350 mg	13	0	428	402	465
8.	PACKAGING & LABELING	13 Individual packs + 13 Consumer packs	0	AS PER AQL	26	0	COMPLIES		

REMARKS: PASS

venkat
(S.VENKATESWARLU) 27/08/22
FINAL LOT TESTING IN-CHARGE

RAJA.R
27/08/2022
(RAJA. R)
Q.A. HEAD

BV - BURST VOLUME

BP - BURST PRESSURE

100 27.11