ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: HBcAb EIA Test Kit

PURCHASE NO.: 37740068A

BATCH NO.: 2008762

EXPIRATION DATE: 2021/11/08

QUANTITY: 11kits

SPECIFICATION	STANDARD	CONCLUSION
Appearance	No visible deposit or floccules in liquid component	Complies
Negative Control	OD＞1.000	Complies
Positive Control	OD＜0.080	Complies
QCO reference	0.500≤QCO/Cutoff value≤1.500	Complies
Compliance rate of negative reference	15/15	Complies
Compliance rate of positive references	15/15	Complies
Precision criteria	CV＜20%	Complies

Quality Department: Hedy zhang

Date: Sep. 23, 2020

Digitally signed by Monastîrschii Viorica
Date: 2021.02.02 11:12:40 EET
Reason: MoldSign Signature
Location: Moldova

REPUBLICA MOLDOVA, mun. CHIȘINĂU
S.R.L.
"SANMEDICO"
IDNO 100
SOCIETATEA CU RĂSPUNDERE LIMITATĂ