

# CERTIFICATE OF COMPLIANCE

According to EMC 2004/108/EC

**Applicant Name** : Power Advance Tech. Co. Ltd  
**Applicant Address** : 2F, 4Cell, Tower C, In.Do Mansion No.48A Zhichun Road,  
Haidian District Beijing, China  
**Manufacturer by** : Power Advance Tech. Co. Ltd  
**Manufacturer's Address** : 2F, 4Cell, Tower C, In.Do Mansion No.48A Zhichun Road,  
Haidian District Beijing, China  
**Product Description** : Universal Tester  
**Model / Part Number** : T500  
**Serial Number(s) Tested** : Prototype  
**Trademark** : PONOVO  
**Date(s) of Testing** : 2008, September 2<sup>th</sup> to 7<sup>th</sup>  
**Test Report Number(s)** : POW-0809-9066-CE

*This is to certify that the product identified above is in compliance with the essential requirements of the following standards:*

**EN 61326: 2006 (Emissions)**

EN55011: 2007, CISPR 11:2003 (Modified) + A1:2004 (Modified), Class A Radiated and Conducted Emissions.

**EN61000-3-2:2006, Harmonic Current Emissions**

**EN61000-3-3: 1995 + A1: 2001+A2:2005, Voltage Fluctuation and Flicker.**

**EN 61326: 2006 (Immunity)**

IEC 61000-4-2 Edition 1.2 (2001), Electrostatic Discharge Immunity Test.

IEC 61000-4-3:2002, Radiated, Radio-Frequency Electromagnetic Field Immunity Test.

IEC 61000-4-4:2004, Electrical Fast Transient/Burst Immunity Test.

IEC 61000-4-5 Edition 1.1:2001, Surge Immunity Test.


IEC 61000-4-6 Edition 2.0 (2003), Immunity to Conducted Disturbances Test.

IEC 61000-4-8 Edition 1.1 (2001), Power Frequency Magnetic Field Immunity Test.

IEC 61000-4-11:2004, Voltage Dips and Short Interruptions Immunity Test.

**Reviewed by:**

**Signed by:**



Paul Chen / QC Manager

**Issued by:**



Worldwide Certification Solutions  
EMC Compliance Management Group  
684 West Maude Avenue Sunnyvale, CA 94085  
650-988-0900(Tel) 650-988-6647(Fax)  
www.ecmg-global.com

**Issued Date:** 2008, September 10

This is the result of tests that were carried out from the submitted product sample(s) in conformity with the specification of the respective standards. The certificate holder has the right to affix the CE-Mark on the inspected product only when the product is completely complying with the required standards.



© 2008 ECMG Worldwide Certification Solutions, Digital Certificate