



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
BS-MDR-099



Product Service

EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class I Devices in sterile condition, with measuring function or reusable surgical instruments)

G11 004593 0012 Rev. 00

Manufacturer:

Shenzhen Antmed Co., Ltd.

18 Jinhui Ave., Pingshan New District
518122 Shenzhen
PEOPLE'S REPUBLIC OF CHINA

SRN Manufacturer - CN-MF-000029659

Authorized Representative:

MedNet EC-REP C IIb GmbH
Borkstrasse 10, 48163 Münster, GERMANY

The Certification Body of TÜV SÜD Product Service GmbH certifies that the manufacturer has established, documented and implemented a quality management system as described in Article 10 (9) of the Regulation (EU) 2017/745 on medical devices. Details on device categories covered by the quality management system are described on the following page(s).

The Report referenced below summarises the result of the assessment and includes reference to relevant CS, harmonized standards and test reports. The conformity assessment has been carried out according to Annex IX Chapter I and III of this regulation with a positive result. As applicable the involvement of the notified body is limited to the aspects relating to:

- establishing, securing and maintaining sterile conditions,
- conformity of the devices with the metrological requirements,
- reuse of the device, in particular cleaning, disinfection, sterilization, maintenance and functional testing and the related instructions for use.

The certified quality management system is subject to periodical surveillance by TÜV SÜD Product Service GmbH. All applicable requirements of the testing and certification regulation of TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:G11 004593 0012 Rev. 00 0012 Rev. 00

Report No.: BJ23081103

Valid from: 2024-02-06

Valid until: 2029-02-05

Christoph Dicks
Head of Certification/Notified Body

Issue date: 2024-02-06



EU Quality Management System Certificate (MDR)

Pursuant to Regulation (EU) 2017/745 on Medical Devices, Annex IX Chapters I and III
(Class I Devices in sterile condition, with measuring function or reusable surgical instruments)

G11 004593 0012 Rev. 00

Classification:

Class I

Device Group:

A020102 - INFUSION AND IRRIGATION SYRINGES, SINGLE-
USE

Device Properties:

MDS 1005.1 - Ethylene Oxide sterilization
MDS 1010 - Devices with a measuring function

Classification:

Class I

Device Group:

A030201 - EXTENSIONS

Device Properties:

MDS 1005.1 - Ethylene Oxide sterilization

Classification:

Class I

Device Group:

C010401 - CARDIAC ANGIOGRAPHY DEVICES

Device Properties:

MDS 1005.1 - Ethylene Oxide sterilization

MDS 1010 - Devices with a measuring function

Classification:

Class I

Device Group:

C900101 - HAEMOSTASIS VALVES

Device Properties:

MDS 1005.1 - Ethylene Oxide sterilization

**The validity of this certificate
depends on conditions and/or
is limited to the following:**

Revision History:

Rev.	Dated	Report	Description
00	2024-02-06	BJ23081103	Initial issuance