


WATO EX-55Pro/65Pro Anesthesia Machine

Service Manual

Intellectual Property Statement

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. (hereinafter called Mindray) owns the intellectual property rights to this product and this manual. This manual may refer to information protected by copyrights or patents and does not convey any license under the patent rights of Mindray, nor the rights of others. Mindray does not assume any liability arising out of any infringements of patents or other rights of third parties.

Mindray intends to maintain the contents of this manual as confidential information. Disclosure of the information in this manual in any manner whatsoever without the written permission of Mindray is strictly forbidden. Release, amendment, reproduction, distribution, rent, adaption and translation of this manual in any manner whatsoever without the written permission of Mindray is strictly forbidden.

mindray、 **MINDRAY** and **WATO** are the registered trademarks or trademarks owned by Mindray in China and other countries. All other trademarks that appear in this manual are used only for editorial purposes without the intention of improperly using them. They are the property of their respective owners.

Contents of this manual are subject to changes without prior notice.

Revision History

This manual has a revision number. This revision number changes whenever the manual is updated due to software or technical specification change. Contents of this manual are subject to change without prior notice. Revision 1.0 is the initial release of the document.

- Revision number: 4.0
- Release time: 2018-09

© Copyright 2016-2018 Shenzhen Mindray Bio-Medical Electronics Co., Ltd. All rights reserved.

The code of one-year service packages:

115-034694-00	one-year maintenance package (0625/Pre-pak)
115-034696-00	one-year expansion package (0625/Pre-pak)
115-034698-00	one-year complete package (0625/Pre-pak)
115-021121-00	one-year maintenance package (0616/0625)
115-021122-00	one-year expansion package (0616/0625)
115-021123-00	one-year complete package (0616/0625)

S/N	P/N	Description	Qty	Usage	Note
1	082-000711-00	O-ring 10X1.8, viton, A70	1	Airway pressure gauge	
2	M6M-010031---	O-ring 27X1.5, silicone, A50, red	2	Valve cover	
3	M6M-010033---	O-ring 20X1.5, silicone, A50, red	2	Valve body	
4	115-001366-00	Flow sensor assembly	1	Breathing circuit	one-year complete package
5	0611-10-45654	Sensor Oxygen MedicelMOX-2	1	Breathing circuit	one-year complete package
6	M6M-010038---	O-ring 23.47X2.95, silicone, A50, red	1	Water collection cup	
7	040-000358-00	Bellows, neoprene, 8 folds	1	Bellows dome assembly	one-year expansion package or one-year complete package
8	M6M-010071---	O-ring 20.29X2.62, silicone, A50, red	1	Top cover 2 assembly	
9	0601-20-69771	PoP-off valve rubber pad (Die MR69771)	1	PoP-off valve rubber pad	
10	049-000243-00	Bellows sealing cushion	1	Bellows dome assembly	
11	0601-20-78840	BYPASS large sealing cushion (Die MR78840)	1	BYPASS large sealing cushion	