

Validation of alternative analysis methods

NF102 - Application to the food industry

Certificate

Certificat No.: BKR 23/07-10/11 Renewal decision dated: 22-09-2023

Expiry date: 07-10-2027

The Company:

SOLABIA

41, rue Delizy 93698 Pantin Cedex - France

Is authorized to affix the NF VALIDATION mark on the alternative analysis method cited below, in accordance with the NF VALIDATION general rules and the certification rules NF102 - Validation of alternative analysis methods (Application to the food industry):

IRIS Salmonella®

Validated for the detection of Salmonella spp.

Technical sheet reference's

IRIS SALMONELLA_V9 (fr)

This decision attests that the alternative analysis method has been assessed by AFNOR Certification and found to conform to the standards cited in page 2/2 and complementary requirements, as specified in the certification reference document. The **principal certified characteristics** are the "analytical performances" as defined in the associated validation study summarized report (sensitivity, relative detection level...), available on the certification dedicated website http://nf-validation.org/en.

This certificate supersedes all previous certificates (last version dated 07-12-2021). This NF VALIDATION certificate, included 2 pages, is valid until 7th October 2027. It is subject to the results obtained upon regular controls carried out by AFNOR Certification. Appropriate decision is made by AFNOR Certification in accordance with the NF VALIDATION general rules and certification rules NF102 - Validation of alternative analysis methods (Application to the food industry).



Managing Director Julien NIZRI

sue dated 25/09/2023

age 1/2



Certificate

Certificat No.: BKR 23/07-10/11 Renewal decision dated: 22-09-2023

Expiry date: 07-10-2027

The alternative analysis method:

IRIS Salmonella®

Validated for the detection of Salmonella spp.

Manufactured by:

SOLABIA

Rue des 40 mines ZAC de Ther Allonne BP 10245 60002 Beauvais - France

Has been certified according to the reference documents and the application scope specified here after:

Validation protocol **NF EN ISO 16140-2** (September 2016): Microbiology of the food chain. Method validation - Part 2: Protocol for the validation of alternative (proprietary) methods against a reference method.

Reference method(s) NF EN ISO 6579-1 (April 2017) and its amendment A1 (March 2020): Microbiology of the food

chain- Horizontal method for the detection, enumeration and serotyping of Salmonella - Part 1:

Detection of Salmonella spp.

Scope All human food products (by performing validation assays on a broad range of foods), pet food and

animal feed (0 à 25g), dry pet food and animal feed (50 à 125g), industrial production environmental

samples.

Restriction(s) None.

Warning None.

Other information Non applicable.

Please send any queries concerning the performances of the certified alternative method to AFNOR Certification.

You may download the validation study summarized report on http://nf-validation.afnor.org/en.

