

**Plaster of Paris Bandage Analysis
Report**

DATE

20/05/2021

PRODUCT NAME: Plaster of Paris Bandage

SIZE : 10CM X 2,70M

LOT NUMBER :120521106

TESTS PERFORMED	TS 4434	CONCLUSION	CONDITIONS OF CONFORMITY
OF THE WRAP SHEET MASS PER UNIT AREA(g/m ²)	MIN 26 g	30,62 gr	A
THE WRAP AND WEFT WIRE WINDING NUMBER OF CLOTH	Cm2 WIRE MUST BE AT LEAST 18 PER	18 wire	A
FEATURE OF THE WRAP SHEET	100% COTTON AND SHOULD BE HYDROPHILE	100% COTTON AND HYDROPHIL	A
PLASTERY MASS PER UNIT AREA OF THE WRAP SHEET	AT LEAST 340 g	436 gr	A
WETTABILITY	WHEN THROWN IN TO THE WATER SHOULD NOT DRY SURFACE	FULL WETTING OCCURRED AFTER 6 SECONDS 30°C WATER	A
SETTING TIME	1-8 MINUTES	2,45 MINUTES	A

CONDITIONS OF CONFORMITY : A:Available NA: None/Available

CONCLUSION: Test results conform to the relevant standard

QUALITY CONTROL OFFICER

GÖZDE DURGİT

MEDİTEKS SAĞLIK HİZMETLERİ
TIBBİ MALZEME TEKSTİL ÜRÜNLERİ
SANDIŞ TİC.A.Ş EDİRNE SÜLOĞLU ŞUBESİ
Organize San. Bölgesi 4.Cd. No:5 Süloğlu-Edirne
Davutpaşa Y.d.No: 613 012 4314
Mersis No: 06 10 0124 314 00016