

Declaration


We,

**RZ Medizintechnik GmbH
Unter Hasslen 20
78532 Tuttlingen
Germany**

herewith confirm following requirements:

- All offered items are manufactured by RZ Medizintechnik GmbH
- Our Instruments posses heat resistance up to a temperature of 200°C and resist sterilization with vapors, up to a temperature of 134°C
- Our instruments are compatible with all common disinfectant solutions for surgical instruments.
- The surface of the instruments is matte.
- The instruments will have following information engraved on the surface:
 - RZ-Logo
 - Reference code
 - Country of origin
 - Stainless
 - CE marking
- The instruments steel has the required composition:
 - C (carbon): 0.30 – 0.45% - Cr (chromium): 15.00 – 18.00% - Mo (molybdenum): 0.90 – 1.30% - Ni (nickel): ≤ 1.00% - Si (silicon): ≤ 1.00% - Mn (manganese): ≤ 1.00% - P (phosphorus): ≤ 0.05%
- The RZ catalogue will be presented at the time of the auction.
- The IFUs are available in English language.
- A product training can be held by a trained specialist, within one month of the request.

Tuttlingen, 11th January. 2024



**Tobias Zubrod
General Manager
RZ Medizintechnik GmbH**

Geschäftsführer
Rainer Zubrod
Tobias Zubrod

Commerzbank AG Tuttlingen
IBAN DE27 6438 0011 0271 5788 00
SWIFT-CODE/BIC: DRESDEFF643

Deutsche Bank Tuttlingen
IBAN DE06 6537 0075 0228 9650 00
SWIFT-CODE/BIC: DEUTDESS653

