

To Whom it May Concern

27 November 2023

### **Certificate of conformity to standards IEC 60118-0**

This certificate confirms that OTICON hearing aids (RUBY 2 BTE, PG10 BTE SP) presented for LP no. MTender ID ocds-b3wdp1-MD-1695909003700 (Acquisition of medical devices (hearing aids), in order to achieve the National Program on hearing aids, for 2024) have been measured according to IEC 60118-0 (Measurement of the performance characteristics of hearing aids) and international standards.

#### **RUBY 2 BTE**

Maximum output: 130 dB;

Maximum gain: 66 dB;

Frequency range: 105 Hz – 7500 Hz

Noise level (omni): 21 dB;

Harmonic distortion coefficient:

500 Hz - <2%

800 Hz - 2%

1600 Hz - <2%

#### **PG10 BTE SP**

Maximum output: 143 dB;

Maximum gain: 83 dB;

Frequency range: 100 Hz - 6500 Hz;

Noise level (omni): 18 dB;

Harmonic distortion coefficient:

500 Hz - 4%

800 Hz - <2%

1600 Hz - <2%

Yours sincerely

Oticon A/S



Claus F. Horneman

Regional Sales Director

