

CERTIFICATE OF CONFORMITY

Customer SA M INTER FARMA
PO MED-X ORDER 01/04/2021
Order # CSM0114642
Glass Type RWB46
Master Specification 07 SPXR 01

We hereby certify that the products delivered are in accordance with your purchase order and Corning Specification. More specifically we certify that minimum thickness measured on the delivered products is specified below.

Lead equivalence thickness has been measured and calculated in accordance with IEC61331-1:2014 and IEC 61331-2:2014 by Public Health England Agency.

Quality Check

Date 14/04/2021

Julie CLEMENT
Production and Quality
Manager



Item	Quantity	Lead equivalence thickness @ kV
10019 CP 1000x800 TH 7.0/8.5	4	LEAD EQUIVALENCE 2.1 MM @ 150 kV
1810N1230/1		
1810N1230/2		
1911N0102/1		
2104N1567T6604/2		