

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

**CB TEST CERTIFICATE**

Product  
Produit

Name and address of the applicant  
Nom et adresse du demandeur

Name and address of the manufacturer  
Nom et adresse du fabricant

Name and address of the factory  
Nom et adresse de l'usine

Note: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>eme</sup> page

Ratings and principal characteristics  
Valeurs nominales et caractéristiques principales

Trademark (if any)  
Marque de fabrique (si elle existe)

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais  
constructeur

Model / Type Ref.  
Ref. De type

Additional information (if necessary may also be  
reported on page 2)  
Les informations complémentaires (si nécessaire,,  
peuvent être indiqués sur la 2<sup>eme</sup> page

A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

As shown in the Test Report Ref. No. which forms part  
of this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

**CERTIFICAT D'ESSAI OC**

Diagnostic Ultrasound System

SIEMENS HEALTHCARE LTD  
26TH FLOOR KINS TOWER8 SEONGNAM-DAERO 331  
BEON-GIL BUNDANG-GU SEONGNAM-SI GYEONGGI-DO  
13558 KOREA

SIEMENS MEDICAL SOLUTIONS USA INC.  
685 EAST MIDDLEFIELD ROAD  
MOUNTAIN VIEW CA 94043 USA

SIEMENS MEDICAL SOLUTIONS USA INC  
SUITE B 2500 MILLBROOK BUFFALO GROVE IL 60089  
USA

Additional Information on page 2

100/115/230 Vac, 15/12/6.5 A, 50/60 Hz



CTF Stage 2

Acuson S1000(10441701)

Additional Information on page 2

IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1, IEC  
60601-1-6(ed.3), IEC 60601-1-6(ed.3);am1, IEC  
60601-2-18(ed.3), IEC 60601-2-37(ed.2), See Page 2

E170920-D15-CB-2-A1 issued on 2016-10-27

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2016-10-31  
Original Issue Date: 2015-12-02

Signature:

Jan-Erik Storgaard



Factories:

JABIL CIRCUIT (SHANGHAI) LTD  
600 TIAN LIN RD SHANGHAI 200233  
CHINA

Additional Information:

Additionally evaluated to EN 60601-1: 2006/ A1:2013/ A12: 2014

The original report was modified to include the following changes/additions:  
Add alternate components, update Applicant information, see test report

Additionally Tested According To:

IEC 62366(ed.1), IEC 62366(ed.1);am1

**Additional information (if necessary)**

**Information complémentaire (si nécessaire)**



- 
- 
- 
- 

UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2016-10-31

Original Issue Date: 2015-12-02

Signature:   
Jan-Erik Storgaard