

Test Report

– Eppendorf PCR clean & sterile –

ep Dualfilter T.I.P.S.[®] 2 - 200 µL



10x 96 Eppendorf Tips
PCR clean / sterile

LOT L2025081

2027-03-28

STERILE R

Order no.: 0030 078.551

Eppendorf AG
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 0030078551

Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

• Human DNA

Method: real-time PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• DNase

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable
(LOD: 1.0 x 10⁻⁶ Kunitz units)

• RNase

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable
(LOD: 1.0 x 10⁻⁹ Kunitz units)

• PCR Inhibition

Method: real-time PCR acc. to SOP 100559/V1
Limit: less than 10 targets amplifiable

• Sterility

Method: Sterilization with SAL 10⁻⁶ by
R: Irradiation according DIN EN ISO 11137:2015* or
EO: Ethylene oxide*

*Please see attached label for the standard method of
sterilization for this product

Test method: USP, Ph. Eur. 8.0 / 2.6.12

• Endotoxin

Method: Kinetic-chromogenic LAL test (Ph. Eur. 8.0 / 2.6.14)
Limit: < 0.001 EU/mL

Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 18.03.22

Testing of the above described lot showed conformity within the limits of detection.

23.03.22
Date of release

ep 2207 641
Registration number

Th. Weigel
Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.