

Cathode Buffer Container 3500 Series

Product No. 4408256
Lot No. 2503082
Date of Manufacture 22MAY2025
Expiration Date 22JUL2026

TEST	SPECIFICATION	RESULT
VISUAL INSPECTION		
Clear, particulate free, viscous solution		Pass
pH TEST Measured at 25.0°C ± 0.3°C		
pH Value	7.6 - 8.2	7.924
CONDUCTIVITY Measured at 25.0°C ± 0.1°C		
Conductivity	1500 - 2000 µS/cm	1768 µS/cm

*For Research Use Only.
Not for use in diagnostic procedures.*

**Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA, USA**

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

Elsa Qirollari
Quality Assurance
Issued 23MAY2025