

Zimmer MedizinSysteme GmbH Junkersstraße 9 89231 Neu-Ulm / Germany Tel.: 0731 / 9761-0 Fax: 0731 / 9761-118 info@zimmer.de www.zimmer.de

Declaration of Conformity

for the electromagnetic stimulator type emFieldPro

We declare under our sole responsibility, that the device

emFieldPro

bears the CE mark

C €₀₁₂₃

according to the Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4) and fulfils the essential requirements of annex I of this directive.

The device belongs to class IIa according to annex IX rule 9 of this directive.

The conformity of the quality management system is certified and controlled by TÜV SÜD Product Service GmbH, Ridlerstraße 65, 80339 Munich, Notified Body Code: 0123. EC certificate no. G1 012889 0044 Rev. 02.

This declaration is valid from 2020-07-15 until 2024-05-26.

Stefan Leinweber Quality Management

Neu-Ulm, 2020-07-15