


## SED

### Extrication/spine immobilization device



The SED extrication device is a first aid apparatus to be used for the extraction of a traumatised patient from a vehicle. It must be used after the application of the cervical collar to maintain immobilisation and head-torso alignment.

#### Specific features

- Color coded belt system to make the application more intuitive and rapid
- PVC coating for an easy sanitation
- The separation of the rigid elements, allows high vertical rigidity and horizontal flexibility
- Equipped with cushion for the nuchal area

#### Technical data

Length	830 mm
Width	900 mm
Maximum thickness <sup>(1)</sup>	25 mm
Overall dimensions wrapped with bag (approx.)	850 x 250 x 120 mm
Belts length	74 ± 2 cm
Materials	PVC, Nylon, PP
Weight without bag	2,60 kg
Weight with bag	2,85 kg
Maximum load capacity	230 kg

<sup>1)</sup> at the hooks

#### Standard equipment

- Transport bag

Class I MD compliant with UE Reg. 2017/745

SED SPENCER EXTRICATION DEVICE



SR00111B

CND Classification V0804

Registration number 102330

NATO stock n° 6515-15-149-8710

Rev.0 (30/11/2020)

UNCHECKED COPY – further revisions will be available on <http://support.spencer.it>  
 Spencer Italia S.r.l. Sala Baganza (PR) Italia Tel. +39.0521.541111 Fax +39.0521.541222  
[www.spencer.it](http://www.spencer.it)