



Technical Specification No Battery aspirator:

- ✓ **Manufacturer: GIMA**
- ✓ **Country of Origin: Italy**
- ✓ **Model: Mini Aspeed Evo Battery Suction 1**
- ✓ **Certification ISO, CE, EN**

Technical Specification	Yes	No	Remarks
<p>Configuration</p> <ul style="list-style-type: none"> ▪ one portable, endowed with a carrying bag; ▪ fall resistant, blows, water and disinfectants; ▪ with a vacuum regulator incorporated; ▪ robust, portable, compact; ▪ electrical operation with a battery incorporated; ▪ continuous mode of operation, based on the built-in battery with autonomy of operation at least 60 min. ▪ power supply 220V, 12V with adapter and built-in battery ▪ free air suction flow 30 L/min; ▪ the maximum pressure will be minimum 600 mmHg; ▪ analogical display or digital of the set suction pressure; ▪ the minimum capacity of the reservoir is 1 L, reusable; ▪ alarm and monitoring system 	<p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p> <p>X</p>		<ul style="list-style-type: none"> • Operating voltage: with battery DC 12 V with power supply 100÷240 V ~ 50/60 Hz - 14 VDC 4,28A • Power consumption: max 50 W • Size: 134 x 378 x h 263 mm Weight: 4.7 kg • Bottle capacity: 1000 ml • Flow: 30 air litres/min • Sound level: 65 dBA at 1 m • Operating cycles: continuous (45 minutes battery autonomy) • MAX vacuum: 0 ± 85 kPa • Internal battery: 12 VDC 4A h lead hermetic • Battery autonomy: about 45 minutes, recharging time approx 6 hours • Norms: <ul style="list-style-type: none"> - MDD 93/42/EEC (Class IIa - annex IX) - EN60601-1(IEC 60601-1) / EN60601-

<p>for the battery status and connection to the power supply;</p> <ul style="list-style-type: none"> it is delivered in a kit with carrying bag, mounting system to the wall of ambulance with cable and power plug at 12V, with minimum 2 reusable silicone tubes of 1,5-2 m in length. Antibacterial filters minimum 5 pieces. Set of Yankauer suction probes – 2 pieces. 			<p>1-2 (IEC 60601-1-2)</p> <p>- UNI EN ISO 10079-1/ UNI CEI EN ISO 14971</p> <p>- Degree of protections against ingress of liquids: IP34d (IEC 529)</p> <p>- Class II- Type B</p> <div> <p>Standard accessories</p> <ul style="list-style-type: none"> • Autoclavable 1 L jar with safety valve • Sterile disposable cannula CH14 • Sterile disposable manual regulator • Disposable silicone liner 1 L • 2 Bacterial hydrophobic filter • Cigarette lighter power cord • Power supply cord - 2 m • Shoulder carrying bag • Multivoltage battery charger • Silicone tubing set Ø 6 x 12 mm • User manual: GB, IT, FR, DE, ES, RO </div>
Certification:			
<ul style="list-style-type: none"> ISO 9001 ISO 10079-1 CE CE Declaration of Conformity 93/42/EEC 	<ul style="list-style-type: none"> X X X X 		<ul style="list-style-type: none"> certificate attached certificate attached certificate attached certificate attached
Other services included in offer			
<ul style="list-style-type: none"> transport to the final customer instalation and put in function training for the technical staff for final customer 	<ul style="list-style-type: none"> X X X X 		<ul style="list-style-type: none"> bidder declaration attached bidder declaration attached bidder declaration attached bidder declaration attached

<input type="checkbox"/> training for the medical staff for final customer	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> warranty for 24 months	X		<input type="checkbox"/> bidder declaration attached
GUARANTEE			
<input type="checkbox"/> 24 months guaranteed from signing the receiving minutes	X		<input type="checkbox"/> bidder declaration attached
SERVICE AND MAINTENANCE			
<input type="checkbox"/> the bidder has technical facilities necessary for the service both for ambulances as well as for medical equipment according to the general guarantee conditions and the user guide of the manufacturer	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> Maximum time for technical intervention – 48 hours.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> Maximum time for remedial measures, in total 72 hours.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> The technical service and current repair will be performed without waiting in a queue.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> Temporary replacement of the equipment has to be ensured according to the periods mentioned above.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> During the guarantee period, at the reasonable request of the user, repair, adjustments and maintenance of medical equipment ,according to the specifications in the manufacturers’ guides, will be done free of charge.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> The spare parts and workmanship are for free, except from the consumables for vehicles established by the manufacturer	X		<input type="checkbox"/> bidder declaration attached
Technical Specification	Yes	No	Remarks

AVAILABILITY OF SPARE PARTS			
<p><input type="checkbox"/> the bidder will ensure the availability of spare parts, accessories and consumables for all offered positions on the market of the Republic of Moldova for at least 10 years from putting into operation, free of charge or against payment, as follows:</p> <ul style="list-style-type: none"> <input type="checkbox"/> spare parts free of charge, including the workmanship for the 36 months of guarantee period <input type="checkbox"/> for the rest of the period – against payment for at least 10 years period is fulfilled. <p><input type="checkbox"/> The bidder will submit the pricelist for the spare parts and the maintenance program to the Ministry of Health, Labour and Social Protection as beneficiary</p>	X		<input type="checkbox"/> bidder declaration attached
	X		<input type="checkbox"/> bidder declaration attached
	X		<input type="checkbox"/> bidder declaration attached
	X		<input type="checkbox"/> bidder declaration attached
GUIDES			
<input type="checkbox"/> Technical Guide and User Guide.	X		<input type="checkbox"/> bidder declaration attached
<input type="checkbox"/> All guides will be available in Romanian /Russian.	X		<input type="checkbox"/> bidder declaration attached