

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME**CB TEST CERTIFICATE**

Product	PCR Thermocycler
Name and address of the applicant	DLAB SCIENTIFIC CO.,LTD. Yu An Road 31, Airport Economic core zone, Shun yi District, Beijing, 101318, P.R. China
Name and address of the manufacturer	DLAB SCIENTIFIC CO.,LTD. Yu An Road 31, Airport Economic core zone, Shun yi District, Beijing, 101318, P.R. China
Name and address of the factory	DLAB SCIENTIFIC CO.,LTD. Yu An Road 31, Airport Economic core zone, Shun yi District, Beijing, 101318, P.R. China
Ratings and principal characteristics	AC 200-240V, 50/60 Hz, 600W, Class I
Trademark (if any)	DLAB
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	TC1000-S, SCI1000-S, 50X4101300 TC1000-G, SCI1000-G, 50X4102300 (X=1-9, X stands for different client)
Additional information (if necessary may also be reported on page 2)	For model differences, refer to the test report.
A sample of the product was tested and found to be in conformity with	IEC 61010-2-010:2019 IEC 61010-1:2010+A1 See Test Report for National Differences
As shown in the Test Report Ref. No. which forms part of this Certificate	CN21IZ62 001

This CB Test Certificate is issued by the National Certification Body



**TÜVRheinland**®

TÜV Rheinland Japan Ltd.  
Global Technology Assessment Center  
4-25-2 Kita-Yamata, Tsuzuki-ku  
Yokohama 224-0021, Japan  
Phone + 81 45 914-3888  
Fax + 81 45 914-3354  
Mail: info@jpn.tuv.com  
Web : www.tuv.com

Date: 2022-04-24

Signature:

Xuhua Liu