

<b>DAC-SpectroMed S.R.L.</b> Chisinau	<b>Declaration of EC-Conformity</b>	File No.	F-PS-10-05
		Rev. No.	2
		Rev. Date	07-16-2019
		Doc.No	STD-DCE-186
		Page	1 / 1

## **DAC-SpectroMed SRL**

**Nicolae Testemitanu str. 37, MD-2025, Chisinau, Republic of Moldova**

hereby declares under its own responsibility that the *in-vitro diagnostics* medical devices:

### **UriControl-DAC**

REF 4062UC16 8 ml + 8 ml

- are classified as not listed in Annex II of the EC Council Directive 98/79/EC from 27<sup>th</sup> October 1998 on in-vitro diagnostic medical devices;
- conform to the relevant provisions of the EC Council Directive 98/79/EC from 27<sup>th</sup> October 1998 on in-vitro diagnostic medical devices.

Harmonized standards applied:

EN ISO 9001:2015

EN ISO 14971:2012

EN ISO 18113-1:2011

EN ISO 18113-2:2011

EN ISO 23640:2011

EN ISO 13612:2002

EN ISO 15223-1:2012

EN ISO 13485:2016

DAC-SpectroMed SRL has a Quality System in place based on EN ISO 9001:2015 and EN ISO 13485:2016, issued by the IQNet&SRAC.

Our Authorized Representative in EU is Qarad B.V., with address at Flight Forum 40, 5657 DB Eindhoven, The Netherlands.



General Manager

Virschi Roman

16.07.2019

Head of Normative-Technical Documentation  
and Certification Department

Emet Natalia

16.07.2019