

Document No.: 812382 Revision: G

Page: 2 of 23

Scheduled replacement may include the following items.

TABLE 1	Interval
Device	
Base air filter	6 months
CCM and CCMB air filter	6 months
Roller Pump air filter	6 months
Electronic Gas System – replace water trap filter	6 months
Electronic Gas System – replace O2 sensor	6 months
Electronic Gas System – replace or recondition the gas system	24 months
Base batteries	24 months
CCMB hard drive – Replacement of hard drive	60 months

PRELIMINARY INSPECTION – Perform a preliminary inspection.

- 6.1 Record Customer/Equipment information.
 - 6.1.1 On the Inspection Report, indicate the type of inspection: Preventive Maintenance (PM), Install, Repair, Update, Remove, or Verification.
 - 6.1.2 Document any customer comments in the Opening Comments section.
 - 6.1.3 Inspect the system for the following broken, damaged, modified, or worn components.
 - 6.1.3.1 Pump/Flow/Centrifugal cables loosely connected cables, loose pins, worn outer sheath, etc.
 - 6.1.3.2 Central Control Monitor (CCM) damage to the screen or cable, loose cable connection to the NIC receptacle, etc.
 - 6.1.3.3 NIC ports damage to any receptacle.
 - 6.1.3.4 Modifications to any system part/component.
 - 6.1.3.5 Missing or broken system parts/components.
 - 6.1.3.6 Document the inspection findings in the Opening Comments.
 - 6.1.4 Record the part numbers and serial numbers for each component of the customer's unit. The part number and serial number information can be found on each component's label.
 - 6.1.4.1 The serial number for the Power Manager Module does not need to be recorded for each system inspection. The Power Manager Module's serial number shall be recorded when the module is replaced with a new or reconditioned one.
 - 6.1.5 Verify and record the software version.
 - 6.1.5.1 Using the current software version table as a guide review all component software versions (in the 'Service and Diagnostics' screen) and update as needed. This is important to ensure that the system will function as intended. Record any changes made in the parts installed section of the record.
 - 6.1.5.2 Record the software version of each component listed in the Service Screen
 - 6.1.6 Also, ensure applicable spare equipment not attached to the system is inspected per this procedure. Use another sheet (i.e. use a copy of page 1 of the inspection form) to document equipment that cannot be listed in the equipment area.