

Certificate of Analysis



Conditioning Reagent 3500 Series

Product No. 4393718
Lot No. 2409499
Date of Manufacture 30SEP2024
Expiration Date 27JUN2025

TEST	SPECIFICATION	RESULT
CONDUCTIVITY TEST Measured and compared against manufacturing specifications		
Conductivity	<= 0.0571 µS/cm	Pass

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system.
Site: Bedford, MA, USA

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at cofarequests@thermofisher.com

Luciana Torcasio
Quality Assurance
Issued 02OCT2024