

# EC Declaration of Conformity

|  |   |
|--|---|
| <b>Manufacturer/<br/>Supplier Information:</b> | <b>BioFire Diagnostics, LLC</b><br>515 Colorow Drive<br>Salt Lake City, Utah 84108, USA<br>Phone: 1-801-736-6354<br><a href="mailto:Regulatory@BioFireDX.com">Regulatory@BioFireDX.com</a><br><a href="http://www.BioFireDX.com">http://www.BioFireDX.com</a> |
|--|---|

We, BioFire Diagnostics, LLC, declare under our sole responsibility, that the product

## **FilmArray® Blood Culture Identification (BCID) Panel (RFIT-ASY-0126, RFIT-ASY-0127)**

meets the provisions of the European Directive 98/79/EC for *In vitro* Diagnostic Medical Devices. The device is classified as a General In Vitro Diagnostic (IVD) Device.

BioFire Diagnostics' quality system is registered to EN ISO 13485:2016.

The following relevant standards have been met:

|   |
|---|
| <b>EN ISO 13485:2016</b><br>Medical devices – Quality Management System – Requirements for regulatory purposes  |
| <b>EN ISO 14971:2012</b><br>Medical devices – Application of risk management to medical devices   |
| <b>EN 13641:2002</b><br>Elimination or reduction of risk of infection related to in vitro diagnostic reagents   |
| <b>EN 62366:2008</b><br>Medical devices-Application of usability engineering to medical devices   |
| <b>EN 13612:2002</b><br>Performance evaluation of in vitro diagnostic medical devices   |
| <b>EN ISO 23640:2015</b><br>In vitro diagnostic medical devices – Evaluation of stability of in vitro diagnostic reagents   |
| <b>EN ISO 15223-1:2016</b><br>Medical Devices – Symbols to be used with medical device labels, labeling and information to be supplied – Part 1: General requirements               |
| <b>EN ISO 18113-1:2011</b><br>In vitro diagnostic medical devices – Information supplied by the manufacturer (labeling) – Part 1: Terms, definition and general requirements        |
| <b>EN ISO 18113-2:2011</b><br>In vitro diagnostic medical devices – Information supplied by the manufacturer (labeling) – Part 2: In vitro diagnostic reagents for professional use |

Technical documentation demonstrating compliance as described in Annex III of the European Directive 98/79/EC is kept by the manufacturer and can be made available by the authorized representative in Europe - QARAD EC-REP BV, Pas 257, B-2440 Geel, Belgium.

The notified body for this product is BSI (Notified body #2797; Say Building, John M. Keynesplein 9, 1066 EP Amsterdam, Netherlands).

Salt Lake City, UT, USA

*(Place and date of issue)*

**Kevin Bourzac**

*Vice President, Regulatory and Clinical Affairs*