

Honeywell

CERTIFICATE OF ANALYSIS

Honeywell Specialty Chemicals Seelze GmbH
Wunstorferstrasse 40
Seelze, 30926
Telefon: +49 5137 999-0
lab.honeywell.com

Brand: Honeywell Fluka
Product: 35354
Lot No: N3490
Production date: 15.Dec.2023
Expiry Date: 29.Nov.2026

Sulfuric acid 0,5 mol/l, volumetric solution Reag. Ph. Eur.

Parameter	Specification	Units	Result	Units
factor	0.999-1.001		1.001	
BAM/EMPA ref.-material	effektiv		93440	
BAM/EMPA Certificate no.	effektiv		U7-0823	
Traceable to NIST SRM	effektiv		723e	

Produced using quality grade "puriss. p.a." tested raw materials.

Test method: Acidimetric titration related to
EMPA/BAM referencematerial (given above)
Product-No 93440: TRIS (Tris(hydroxymethyl)aminoethane).

The factor (titer) was measured at 20°C.

BAM: Federal Institute for Materials Research and Testing
Germany
EMPA: Swiss Federal Laboratories for Materials Testing and Research
NIST: National Institute of Standards and Technology
USA

QC Release Date: 08.Jan.2024

The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed
flasks/ packages.

We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.

Particular properties of the products or the suitability for a particular area of application are not assured.

We guarantee a proper quality within our General Conditions of Sales.



Honeywell
Quality Management

This document was produced electronically and is valid without a signature