



ANTİMİKROP AR-GE VE BİYOSİDAL ANALİZ MERKEZİ



TEST RESULTS (Virucidal Test)

RESULT

The method verified test is valid only when it meets the following verification requirements.

Conditions	Comment
1.1. Virus test suspension should show 4 log difference in virus titer.	Virus titer is enough to show 4 log difference.
1.2. The cytotoxicity of the product should not prevent 4 log difference in virus titer.	Product toxicity does not prevent 4 log difference.
1.3. Interference control test result should not show more than 1 log difference in virus titer compared to virus control.	Interference control is convenient.
1.4. The virus titer difference should be ≤ 0.5 log as a result of the activity suppression test.	Reconstitution with ice cold medium does not prevent virus reproduction.
1.5. The reference inactivator >4 log virucidal effect should be appropriate in the >60 minute test.	Formaldehyde used as reference inactivator showed >4 log virucidal effect in >60 minutes.

Based on TS EN 14476 + A1: 2019-09 standard; the batch of **DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES** when tested at RTU (ready-to-use) possesses virucidal activity in 120 seconds at 20°C under clean conditions for the reference strain of Murine Norovirus virus.

DATE: 26.06.2020

ANKARA

Manager, Microbiological Analysis Laboratory Uzman Bio. Fulya ERLER	Responsible for QMS Nihan Saran EROĞLU	Responsible Manager Sorumlu Yönetici Prof. Dr. Murat ERTÜRK		
ANTİMİKROP BİYOSİDAL ANALİZ MERKEZİ AR-GE / İHAR. VE DAE / İHAR. KONYA SAN. TİC. LTD. ŞTİ 15010 Akıncı Mah. Süleyman İsmailoğlu Caddesi No: 3/1 Çankaya / ANKARA Hesap No: 070 037 5715 - Mersis No: 0836015715000000				
DOCUMENT NUMBER PR-13-FR-45-03	RELEASE DATE 09.07.2019	REVISION DATE 13.04.2020	REVISION DATE 02	PAGE 3/3



Deliklikaya Mah. Kısıklı Cad.
No: 29/05A (PK: 34555)
Amasurköy / İstanbul
B Çekmece Y.D. 03-0369298
Tic. Sic. No: 287509
Mersis No: 0836036929800037

TEST RESULTS (Virucidal Test)

Test	Density	Cytotoxicity (log)	After 1-2-5-15-30-60 Minutes Log TCID50						After 2 Minutes > 4 lg Reduction
			1	2	5	15	30	60	
DEEP FRESH ANTIVIRAL ALKOLLÜ MENDİL/DEEP FRESH ANTIVIRAL ALCOHOL WIPES	RTU (97%)	0,5	-	0,5	-	-	-	-	≥8,67
Formaldehyde	1,4 %	0,5	-	-	-	-	5,83	3,83	-
Virus Control	-	-	-	9,17	-	-	-	-	-

[illegible]

<u>DOCUMENT NUMBER</u> PR-13-FR-45-03	<u>RELEASE DATE</u> 09.07.2019	<u>REVISION DATE</u> 13.04.2020	<u>REVISION DATE</u> 02	<u>PAGE</u> 2/3
--	-----------------------------------	------------------------------------	----------------------------	--------------------

TEST RESULTS (Virucidal Test)

RESULT



The method verified test is valid only when it meets the following verification requirements.

Conditions	Comment
1.1. Virus test suspension should show 4 log difference in virus titer.	Virus titer is enough to show 4 log difference.
1.2. The cytotoxicity of the product should not prevent 4 log difference in virus titer.	Product toxicity does not prevent 4 log difference.
1.3. Interference control test result should not show more than 1 log difference in virus titer compared to virus control.	Interference control is convenient.
1.4. The virus titer difference should be ≤ 0.5 log as a result of the activity suppression test.	Reconstitution with ice cold medium does not prevent virus reproduction.
1.5. The reference inactivator >4 log virucidal effect should be appropriate in the >60 minute test.	Formaldehyde used as reference inactivator showed >4 log virucidal effect in >60 minutes.

*Based on TS EN 14476 + A1: 2019-09 standard; the batch of **DEEP FRESH ANTIVIRAL ALKOLLÜ MENDİL/DEEP FRESH ANTIVIRAL ALCOHOL WIPES** when tested at RTU (ready-to-use) possesses virucidal activity in 120 seconds at 20 °C under clean conditions for the reference strain of Adenovirus Type 5 virus.*

DATE: 26.06.2020

ANKARA

Manager, Microbiological Analysis Laboratory Uzman Bio. Fulya PDK 	Responsible for QMS Nihon Seisan Erciye Nf-	Responsible Manager Sorumlu Yönetici Prof. Dr. Murat ERTÜRK  ANTIMİKROBİYAL ANTİMİKROBİYAL MAD. LAB. AR-GE MÜH. VE ORT. HİZ. KİMYA SAN. TİC. LTD. Şİ. Nispetiye Akad. Mah. Süleyman Hacıoğlu Bulvarı Kat: No 33 Çankaya / ANKARA Tel: 0312 360 00 00 / 0312 360 00 00 / 0312 360 00 00
--	--	---

<u>DOCUMENT NUMBER</u> PR-13-FR-45-03	<u>RELEASE DATE</u> 09.07.2019	<u>REVISION DATE</u> 13.04.2020	<u>REVISION DATE</u> 02	<u>PAGE</u> 3/3
--	-----------------------------------	------------------------------------	----------------------------	--------------------

**ANTİMİKROP AR-GE VE BİYOSİDAL ANALİZ MERKEZİ****TEST RESULTS (Virucidal Test)**

TESTING LABORATORY
ISO/IEC 17025:2017
NAC-032-TR

Test	Density	Cytotoxicity (log)	After 1-2-5-15-30-60 Minutes Log TCID50						After 2 Minutes > 4 lg Reduction
			1	2	5	15	30	60	
DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES	RTU (97%)	0,5	-	4,17	-	-	-	-	≥5,00
Formaldehyde	1,4 %	0,5	-	-	-	-	5,83	3,83	-
Virus Control	-	-	-	9,17	-	-	-	-	-

Test	Density	Contact Time	Dilution (Log) ^a						
			-2	-3	-4	-5	-6	-7	-8
DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES	RTU (97%)	2 Minute	444444	321020	000000	000000	000000	000000	000000
DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES Cytotoxicity	RTU (97%)	-	000000	000000	000000	000000	000000	000000	000000
Formaldehyde	1,4 %	30 Minute	444444	444444	433333	020001	000000	000000	000000
		60 Minute	444444	001001	000000	000000	000000	000000	000000
Formaldehyde Cytotoxicity	1,4 %	-	000000	000000	000000	000000	000000	000000	000000
Virus Control	-	2 Minute	444444	444444	444444	444444	444444	444444	432100

DOCUMENT
NUMBER
PR-13-FR-45-03

RELEASE DATE
09.07.2019

REVISION DATE
13.04.2020

REVISION DATE
02

PAGE
2/3



Beliklikova Mah. Kısıklı Cad.
No: 23/23A / FK: 34055
Antakya-Öy / Hatay
B.Çekmece V.D: 0340389286
Tic. Sic. No: 687509
Mersis No: 00340389286000037

ANTİMİKROP ANTİMİKROBİYAL MAD. LAB.
AR-GE MÜHÜRÜ DOKÜMAN KONTROLÜ LTD ŞTİ
Nispetiye Akademi Sok. No: 10 Kat: 3 No: 3011
Cennetköy / Beşiktaş / İstanbul
Hüvi Yö. No: 070 031 8715 Sicil No: 0070031871500013

	ANTİMİKROP AR-GE VE BİYOSİDAL ANALİZ MERKEZİ		 TESTING LABORATORY ISO/IEC 17025:2017 NAC-013-TL
	TEST RESULTS (Virucidal Test)		

Test Name: Evaluation of virus killing effectiveness of Bi-20-0085 coded product according to TS EN 14476 A1 (September 2019)

Product Name: DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES **Product Code:** Bi-20-0085 (SUDUM-V-0042)


Test Standard: TS EN 14476+A1:2019-09

Test Date: 22.06.2020

Test Conditions:	Instruction to Use	RTU-Modified Method (97%) (Ready to use)
	Diluent	-
	Test Density	100 % Concentration
	Appearance When Reconstituted	-
	Contact Time	2 Minute
	Test Temperature	20 °C
	Interfering Substance	Clean Conditions: 0.3 g / l Bovine Albumin Solution
	Appearance of the Test Item	-
	Incubation Temperature	37 °C
	Neutralization Method	Dilution Neutralization
	Test Organism	Adenovirus Type 5 (Adenoid 75)
	Test Cell	VERO cell line

Analyst **Name/Surname:** Murat ERTÜRK **Signature :** 

Figure 1. The virucidal effect of DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES against Adenovirus Type 5 in clean conditions



DOCUMENT NUMBER
PR-13-FR-45-03

RELEASE DATE
09.07.2019

REVISION DATE
13.04.2020

REVISION DATE
02

PAGE
1/3

TEST RESULTS (Virucidal Test)

RESULT

The method verified test is valid only when it meets the following verification requirements.

Conditions	Comment
1.1. Virus test suspension should show 4 log difference in virus titer.	Virus titer is enough to show 4 log difference.
1.2. The cytotoxicity of the product should not prevent 4 log difference in virus titer.	Product toxicity does not prevent 4 log difference.
1.3. Interference control test result should not show more than 1 log difference in virus titer compared to virus control.	Interference control is convenient.
1.4. The virus titer difference should be ≤ 0.5 log as a result of the activity suppression test.	Reconstitution with ice cold medium does not prevent virus reproduction.
1.5. The reference inactivator >4 log virucidal effect should be appropriate in the >60 minute test.	Formaldehyde used as reference inactivator showed >4 log virucidal effect in >60 minutes.

Based on TS EN 14476 + A1: 2019-09 standard; the batch of DEEP FRESH ANTIVIRAL ALKOLLÜ MENDİL/DEEP FRESH ANTIVIRAL ALCOHOL WIPES when tested at RTU (ready-to-use) possesses virucidal activity in 120 seconds at 20 °C under clean conditions for the reference strain of Poliovirus Type 1 virus.

DATE: 26.06.2020

ANKARA

Manager, Microbiological Analysis Laboratory Uzman Bio. Fulya PAK N.Ş.	Responsible for QMS Nihon Seven EROGLU N.Ş.	Responsible Manager Sorumlu Yönetici Prof. Dr. Mustafa ERGÜRK
--	--	--

ANTIMIKROP ANTIBIOTIKSIAL NAD LAB
AR-GE MUH. /DAN INS JUMTA SANKIT LTD ST
Nasuh Rasyid Mah. Bulevar Antikrop Antibiotik Cus. No.37
Cukrova / ANKARA
Halk St. No.6700218715 Mersis No:067002187150001

<u>DOCUMENT NUMBER</u> PR-13-FR-45-03	<u>RELEASE DATE</u> 09.07.2019	<u>REVISION DATE</u> 13.04.2020	<u>REVISION DATE</u> 02	<u>PAGE</u> 3/3
--	-----------------------------------	------------------------------------	----------------------------	--------------------

**ANTİMİKROP AR-GE VE BİYOSİDAL ANALİZ MERKEZİ****TEST RESULTS (Virucidal Test)**

FESTİM LABORATUVAR
ISO/IEC 17025:2017
NAC-002-15

Test	Density	Cytotoxicity (log)	After 1-2-5-15-30-60 Minutes Log TCID50						After 2 Minutes > 4 lg Reduction
			1	2	5	15	30	60	
DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES	RTU (97%)	0,5	-	4,67	-	-	-	-	≥4,16
Formaldehyde	1,4 %	0,5	-	-	-	-	5,83	3,83	-
Virus Control	-	-	-	8,83	-	-	-	-	-

Test	Density	Contact Time	Dilution (Log) ^a						
			-2	-3	-4	-5	-6	-7	-8
DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES	RTU (97%)	2 Minute	444444	444344	02000	000000	000000	000000	000000
DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES Cytotoxicity	RTU (97%)	-	000000	000000	000000	000000	000000	000000	000000
Formaldehyde	1,4 %	30 Minute	444444	444444	433333	020001	000000	000000	000000
		60 Minute	444444	001001	000000	000000	000000	000000	000000
Formaldehyde Cytotoxicity	1,4 %	-	000000	000000	000000	000000	000000	000000	000000
Virus Control	-	2 Minute	444444	444444	444444	444444	444444	442444	400004

DOCUMENT
NUMBER
PR-13-FR-45-03

RELEASE DATE
09.07.2019

REVISION DATE
13.04.2020

REVISION DATE
02

PAGE
2/3



Deliçikaya Mah. Kısıklı Cad.
No: 23/25A (PK: 34855)
Ataşehir / İstanbul
B Çekmece V.D. : 0340300000
Tic. Sic. No: 287209
Mersis No: 0834030000000007

ANTİMİKROP ANTİMİKROBİYAL MAD. LAB.
AR-GE - MÜH. VE DEN. İŞL. KURUMU LTD. ŞTİ.
Nispetiye Mah. Sütlüce / Nispetiye / Beşiktaş / İstanbul / Türkiye
Çekmece / İstanbul
Hizmet No: 010 031 8715 / Mersis No: 0070031571500013


 ANTİMİKROP AR-GE VE BİYOSİDAL ANALİZ MERKEZİ			
TEST RESULTS (Virucidal Test)			
Test Name:	Evaluation of virus killing effectiveness of Bİ-20-0085 coded product according to TS EN 14476 A1 (September 2019)		
Product Name:	DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES	Product Code:	Bİ-20-0085 (SUDUM-V-0042)
Test Standard:	TS EN 14476+A1:2019-09		
Test Date:	22.06.2020		
Test Conditions:	Instruction to Use	RTU-Modified Method (97%) (Ready to use)	
	Diluent	-	
	Test Density	100 % Concentration	
	Appearance When Reconstituted	-	
	Contact Time	2 Minute	
	Test Temperature	20 °C	
	Interfering Substance	Clean Conditions: 0.3 g / l Bovine Albumin Solution	
	Appearance of the Test Item	-	
	Incubation Temperature	37 °C	
	Neutralization Method	Dilution Neutralization	
	Test Organism	Poliovirus Type 1 (LSc 2ab strain)	
	Test Cell	VERO cell line	
Analyst	Name/Surname: Murat ERTÜRK		Signature : 

Figure 1. The virucidal effect of DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES against Poliovirus Tip 1 in clean conditions



DOCUMENT NUMBER PR-13-FR-45-03	RELEASE DATE 09.07.2019	REVISION DATE 13.04.2020	REVISION DATE 02	PAGE 1/3
--	-----------------------------------	------------------------------------	----------------------------	--------------------

**ANTİMİKROP ARGE VE BİYOSİDAL ANALİZ MERKEZİ****MICROBIOLOGICAL ANALYSIS RESULT REPORT**

TEKİŞİM LABORATUVARI
ISO/IEC 17025:2017
NAC-512-TL

APPENDIX 1: RESULTS

ANALYSIS TYPE AND METHOD	MICROORGANISM USED	RESULT	BACTERICIDAL EFFECT (DECREASE %)	CONTACT TIME
BACTERICIDAL ANALYSIS	<i>P. aeruginosa</i>	-	-	-
	<i>E. hirae</i>	-	-	-
	<i>S. aureus</i>	-	-	-
	<i>E. coli</i>	-	-	-
	<i>E. coli</i> K12	-	-	-
	<i>B. subtilis</i>	-	-	-

	MICROORGANISM USED	RESULT	VIRUCIDAL EFFECT (DECREASE %)	CONTACT TIME
VIRUCIDAL ANALYSIS	Poliovirus Type 1 (LSc 2ab strain) *	4,16 LOG	>%99,99	2 MIN.
	Adenovirus Type 5 (Adenoid 75 strain) *	5,00 LOG	>%99,999	2 MIN.
	M. Norovirus (S99 Berlin strain) *	8,67 LOG	>%99,999999	2 MIN.

	MICROORGANISM USED	RESULT	FUNGUCIDAL EFFECT (DECREASE %)	CONTACT TIME
FUNGICIDAL ANALYSIS	<i>C. albicans</i>	-	-	-
	<i>A. brasiliensis</i>	-	-	-

SUMMARY OF ANALYSIS METHOD	Virucidal tests with DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES were performed according to TS EN 14476 + A1: 2019-09 standards against Poliovirus Type 1, Adenovirus Type 5 and Murine Norovirus at 20 °C for a 2 minutes contact time under clean conditions (0.3 g / l bovine albumin solution).
-----------------------------------	---

This report cannot be partially copied or reproduced without the written permission of the laboratory.
Unsigned reports are invalid
The tests marked with (*) are within the scope of accreditation

DOCUMENT NUMBER PR-13-FR-41-02	RELEASE DATE 01.10.2019	REVISION DATE 06.06.2020	REVISION NO 02	PAGE 2/2
--	-----------------------------------	------------------------------------	--------------------------	--------------------



Delikaya Mah. Kışık Cad.
No: 23/254 (FR 34555)
Arnavutköy - İstanbul
B.Çekmece V.D. : 0340389296
Tic. Sic. No: 087509
Mersis No: 0834038929600037

ANTİMİKROP ANTİMİKROBİYAL MAD. LAB.
AR-GE - MÜH. VE DEN. HİZ. KURUMU TİC. LTD. ŞTİ.
Nispetiye Akademi Sok. No: 37/1 Kat: 3/1
Marmara No: 020 031 0115 Mersis No: 0070031571500013

**ANTİMİKROP ARGE VE BİYOSİDAL ANALİZ MERKEZİ****MICROBIOLOGICAL ANALYSIS RESULT REPORT**

TESTING LABORATORY
ISO/IEC 17025:2017
NAC-015-TL

REGISTRATION NUMBER	R-20-0085
REPORT DATE	26.06.2020
SAMPLE REGISTRATION NUMBER	BI-20-0085 (SUDUM-V-0042)
NAME OF THE PRODUCT	DEEP FRESH ANTİVİRAL ALKOLLÜ MENDİL/DEEP FRESH ANTİVİRAL ALCOHOL WIPES
PRODUCT / LICENSE OWNER	AKSAN KOZMETİK SAN. VE TİC. A. Ş.
ACTIVE SUBSTANCE COMPOSITION	EtOH %75 (v/v)
FORMULATION	-
PACKING MATERIAL	Wet Wipes
SAMPLE ARRIVAL DATE	22.06.2020
INSTITUTION SENDING THE SAMPLE	AKSAN KOZMETİK SAN. VE TİC. A. Ş.
TEST PURPOSE, THE SEAL STATUS and THE QUANTITY	BIOCIDAL ANALYSIS- UNSEALED
SAMPLE PRODUCTION PLACE ADDRESS	AKSAN KOZMETİK SAN. VE TİC. A. Ş. Delikkaya Mah. Kısıklı Cd. No:23/25A Arnavutköy/İSTANBUL
PRODUCTION AND EXPIRATION DATE	19.06.2020/19.06.2022
SAMPLE CHARGER / SERIAL NUMBER	-
ANALYSIS START AND END DATE	22.06.2020-26.06.2020
ANALYSIS METHOD	VIRUCIDAL ANALYSIS
ANALYSIS RESULTS	Appendix 1/ PR-13-FR-45-03 Virucidal Test Result Form

Date:

26.06.2020

Analyst

Murat ERTÜRK

Date: 26.06.2020

Report Prepared by

Gamze DİNGER CENGİZ

Date:

26.06.2020

Responsible Manager

Sorumlu Yönetici
Prof. Dr. Murat ERTÜRKDOCUMENT
NUMBER
PR-13-FR-41-02RELEASE DATE
01.10.2019REVISION DATE
06.06.2020REVISION NO
02PAGE
1/2

Hizmet No: 070/021.8715 Menü No: 007003182/000013

AKSANKOZMETİK
SAN. VE TİC. A.Ş.Delikkaya Mah. Kısıklı Cad.
No: 23/25A / Fk: 34850
Arnavutköy / İstanbul
Bilgi Teknoloji ve İletişim Kurumu
Tic. Sic. No: 287509
Mersis No: 080300388298000037



ANTİMİKROP AR-GE VE BİYOSİDAL ANALİZ MERKEZİ



TEST RESULTS (Virucidal Test)

RESULT

The method verified test is valid only when it meets the following verification requirements.

Conditions	Comment
1.1. Virus test suspension should show 4 log difference in virus titer.	Virus titer is enough to show 4 log difference.
1.2. The cytotoxicity of the product should not prevent 4 log difference in virus titer.	Product toxicity does not prevent 4 log difference.
1.3. Interference control test result should not show more than 1 log difference in virus titer compared to virus control.	Interference control is convenient.
1.4. The virus titer difference should be ≤ 0.5 log as a result of the activity suppression test.	Reconstitution with ice cold medium does not prevent virus reproduction.
1.5. The reference inactivator >4 log virucidal effect should be appropriate in the >60 minute test.	Formaldehyde used as reference inactivator showed >4 log virucidal effect in >60 minutes.

Based on TS EN 14476 + A1: 2019-09 standard; the batch of **DEEP FRESH ANTIVIRAL ALCOHOL WIPES** when tested at RTU (ready-to-use) possesses virucidal activity in 2 and 15 minutes at 20 °C under clean conditions for the reference strain of COVID-19 (SARS-CoV-2) (Clinical Isolate) virus.

DATE: 02.09.2020

ANKARA

Manager, Microbiological Analysis Laboratory Uzman Bio. Fulya PAK	Responsible for QMS Nihan SEVEN ERGÖLÜ	Responsible Manager Sorumlu Yönetici Prof. Dr. Murat ERTÜRK ANTİMİKROP ANTIMİKROBYAL MADDELER AR-GE MÜHÜR VE DAKİLMIZ KİMYA SAN. TİC. LTD. ŞTİ. Yasın Akar Mah. Süleyman Paşa Abdullahoğlu Cad. No: 37/1 Bağcı - Çankaya / ANKARA Tic. Sic. No: 070 031 8715 Mersis No: 0670031871500013
---	---	---

DOCUMENT
NUMBER
PR-13-FR-45-03

RELEASE DATE
09.07.2019

REVISION DATE
13.04.2020

REVISION NO
02

PAGE
3/3



Deliklikaya Mah. Kışkır Cad.
No: 23/25A (PK: 34655)
Arnavutköy / İstanbul
B Çakıncı V.D. 0340368206
Tic. Sic. No: 687508
Mersis No: 0034036829600037

**ANTİMİKROP AR-GE VE BİYOSİDAL ANALİZ MERKEZİ**

TESTING LABORATORY
ISO/IEC 17025:2017
NAC-015-TL

TEST RESULTS (Virucidal Test)

Test	Density	Cytotoxicity (log)	After 1-2-5-10-15-20 minutes Log TCID50						After 2 and 15 minutes > 4 lg Reduction
			1	2	5	10	15	20	
DEEP FRESH ANTIVIRAL ALCOHOL WIPES	RTU (97%)	3,5	-	1,33	-	-	1,17	-	4,84 (2 Min.) 5,00 (15 Min.)
Virus Control	-	-	-	-	-	-	6,17	-	

Test	Density	Contact Time	Dilution (Log) ^a						
			-2	-3	-4	-5	-6	-7	-8
DEEP FRESH ANTIVIRAL ALCOHOL WIPES	RTU (97%)	2 Minutes	043444	000000	000000	000000	000000	000000	000000
		15 Minutes	434004	000000	000000	000000	000000	000000	000000
DEEP FRESH ANTIVIRAL ALCOHOL WIPES Cytotoxicity	RTU (97%)		443444	000000	000000	000000	000000	000000	000000
Virus Control		15 Minutes	444444	444444	444444	044404	033203	000000	000000

ANTİMİKROP ANTİMİKROBİYAL MAD.LAB.
AR-GE MÜH. VE DAN. MÜD. KİMYA SAN. TİC. LTD. STİ.
Nasuh Akar Mah. Süleyman Paşa Bulvarı No: 227, Kat: 1
Beşiktaş - Çarşı - Beşiktaş / İstanbul
HITİT MÜD. No: 070 031 8715 Mersis No: 007603187150001

**DOCUMENT
NUMBER**
PR-13-FR-45-03

RELEASE DATE
09.07.2019

REVISION DATE
13.04.2020

REVISION NO
02

PAGE
2/3



AKSAN VETİK A.Ş.

Değirlikaya Mah. Kısıklı Cad.
No: 23/25A (PK. 34566)
Antayutköy / Beşiktaş
B. Çekmece / M.D. : 0312 368 296
Tic. Sic. No: 267500
Mersis No: 00340380296000037



ANTİMİKROP AR-GE VE BİYOSİDAL ANALİZ MERKEZİ



TEST RESULTS (Virucidal Test)

Test Name: Evaluation of virus killing effectiveness of BI-20-0163 coded product according to TS EN 14476 A1 (September 2019)

Product Name: DEEP FRESH ANTIVIRAL ALCOHOL WIPES **Product Code:** BI-20-0163 (SUDUM-CV-0072)


Test Standard: TS EN 14476+A1:2019-09

Test Date: 28.08.2020

Test Conditions:	Instruction to Use	Ready to use
	Diluent	-
	Test Density	97 % Concentration
	Appearance When Reconstituted	-
	Contact Time	2-15 Minute
	Test Temperature	20 °C
	Interfering Substance	Clean Conditions: 0.3 g / l Bovine Albumin Solution
	Appearance of the Test Item	-
	Incubation Temperature	37 °C
	Neutralization Method	Dilution Neutralisation
	Test Organism	COVID-19 (SARS-CoV-2) (Clinical Isolate)
	Test Cell	VERO E6 cell line

Analyst **Name/Surname:** Murat ERTÜRK **Signature :**

Figure 1. The virucidal effect of DEEP FRESH ANTIVIRAL ALCOHOL WIPES against COVID-19 (SARS-CoV-2) (Clinical Isolate) in clean conditions



DOCUMENT NUMBER
PR-13-FR-45-03

RELEASE DATE
09.07.2019

REVISION DATE
13.04.2020

REVISION NO
02

PAGE
1/3



Değirliçkaya Mah. Kirekli Cad.
No: 23/25A (PK: 34555)
Anadoluköy / Beşiktaş / İstanbul
B. Çekmece V.D.: 0340368296
Tic. Sic. No: 687509
Mersis No: 0834033929600037

ANTİMİKROP ANTİMİKROBİYAL MAD. LAB.
AR-GE MÜH. VE SAN. ÜZ. KİMYA SAN. TİC. LTD. ŞTİ.
Nispetiye Mah. Nispetiye Sok. No: 1/1 Kat: 3/3
Etiler - Beşiktaş / İstanbul
Tic. Sic. No: 070 031 8715 Mersis No: 0870031871500013



ANTİMİKROP ARGE VE BİYOSİDAL ANALİZ MERKEZİ



MICROBIOLOGICAL ANALYSIS RESULT REPORT

TESTİNG LABORATORY
ISO/IEC 17025:2017
NAC-192-21

APPENDIX 1: RESULTS

ANALYSIS TYPE AND METHOD	MICROORGANISM USED	RESULT	BACTERICIDAL EFFECT (DECREASE %)	CONTACT TIME
BACTERICIDAL ANALYSIS	<i>P. aeruginosa</i>	-	-	-
	<i>E. hirae</i>	-	-	-
	<i>S. aureus</i>	-	-	-
	<i>E. coli</i>	-	-	-
	<i>E. coli</i> K12	-	-	-
	<i>B. subtilis</i>	-	-	-
VIRUCIDAL ANALYSIS	MICROORGANISM USED	RESULT	VIRUCIDAL EFFECT (DECREASE %)	CONTACT TIME
	Poliovirus Type 1 (LSc 2ab strain) *	-	-	-
	Adenovirus Type 5 (Adenoid 75 strain) *	-	-	-
	M. Norovirus (S99 Berlin strain) *	-	-	-
	COVID-19 (SARS-CoV-2) (Clinical isolate)	4,84 LOG	>99,99%	2 MIN
		5,00 LOG	>99,99%	15 MIN
FUNGICIDAL ANALYSIS	MICROORGANISM USED	RESULT	FUNGUCIDAL EFFECT (DECREASE %)	CONTACT TIME
	<i>C. albicans</i>	-	-	-
	<i>A. brasiliensis</i>	-	-	-
SUMMARY OF ANALYSIS METHOD	Virucidal tests with DEEP FRESH ANTIVIRAL ALCOHOL WIPES were performed according to TS EN 14476 + A1: 2019-09 standards against COVID-19 (SARS-CoV-2) (Clinical isolate) at 20 °C for a 2 and 15 minutes contact time under clean conditions (0.3 g / l bovine albumin solution).			
This report cannot be partially copied or reproduced without the written permission of the laboratory Unsigned reports are invalid. The tests marked with (*) are within the scope of accreditation.				
DOCUMENT NUMBER PR-13-FR-41-02	RELEASE DATE 01.10.2019	REVISION DATE 06.06.2020	REVISION NO 02	PAGE 2/2



Deliklikaya Mah. Kısıklı Cad.
No: 29-25A (PK: 34555)
Ambarlıköy / İstanbul
B Çekmece Y.D. +0340389296
Tic. Sic. No: 687309
Mersis No: 0034038929600037

ANTİMİKROP ANTİMİKROBİYAL MAD. LAB.
AR-GE MÜH. VE DAN. HİZ. KİMYA SAN. TİC. LTD. ŞTİ.
Nispetiye Mah. Süleyman Hacı Paşa Cad. No: 37/1
Bakır - Çarşı / Beşiktaş / İstanbul
Hizmet V.D. No: 070031 RPA Mersis No: 087003107150013



MICROBIOLOGICAL ANALYSIS RESULT REPORT

TESTING LABORATORY
ISO/IEC 17025:2017
RMC-013-11

REGISTRATION NUMBER	R-20-0163
REPORT DATE	02.09.2020
SAMPLE REGISTRATION NUMBER	BI-20-0163 (SUDUM-CV-0072)
NAME OF THE PRODUCT PRODUCT / LICENSE OWNER	DEEP FRESH ANTIVIRAL ALCOHOL WIPES AKSAN KOZMETİK SAN. VE TİC. A.Ş.
ACTIVE SUBSTANCE COMPOSITION	EtOH (75 % v/v)
FORMULATION	-
PACKING MATERIAL	Wet Wipes
SAMPLE ARRIVAL DATE	13.08.2020
INSTITUTION SENDING THE SAMPLE TEST PURPOSE, THE SEAL STATUS and THE QUANTITY	AKSAN KOZMETİK SAN. VE TİC. A.Ş. Biocidal Analysis-Unsealed
SAMPLE PRODUCTION PLACE ADDRESS	AKSAN KOZMETİK SAN. VE TİC. A.Ş. Delikkaya Mah. Kısıklı Cad. No:23/25A Arnavutköy/İSTANBUL
PRODUCTION AND EXPIRATION DATE	07.08.2020-07.08.2022
SAMPLE CHARGER / SERIAL NUMBER	-
ANALYSIS START AND END DATE	28.08.2020-01.09.2020
ANALYSIS METHOD	Virucidal Analysis
ANALYSIS RESULTS	Appendix 1/ PR-13-FR-45-03 Virucidal Test Result Form

Date:

02.09.2020

Analyst

Murat ERTÜRK

ARM

Date: 02.09.2020

Report Prepared by

Gamze Dincer Cerpiç



Date:

02.08.2020

Responsible Manager

Sorumlu Yönetici

Prof. Dr. Murat ERTÜRK

ANTİMİKROB ANTİMİKROBİTAL MAD. LAB.
AR-GE MÜH. VE SAN. İZ. KİMYA SAN. TİC. LTD. STİ.
Nispetiye Akar Mah. Süleyman Mac. Abdullahoğlu Cad. No 37/1
Balgat - Çankaya / ANKARA
HİTİT V. D. Nispetiye

Balgat - Çankaya / ANKARA
HITIT V. D. No: 070 031 9715 Mersis No: 0670031971500000

DATE _____

<u>DOCUMENT NUMBER</u> PR-13-FR-41-02	<u>RELEASE DATE</u> 01.10.2019	<u>REVISION DATE</u> 06.06.2020	<u>REVISION NO</u> 02	<u>PAGE</u> 1/2
--	-----------------------------------	------------------------------------	--------------------------	--------------------

aksan
KÜZMİLLİ
SANAYİTİ A.Ş.

Delilikaya Mah. Kışık Cad.
No: 22/5A (F3, 34555)
Amasuntköy / İstanbul
B Çekimece Y.D : 03-0389296
Tic. Sic.No: 887509
Mersis No: 0004038929600037