

## Certificate of Compliance

<b>Product Description</b>	:	<b>Cell Culture Slide Hybridwell™ (DLux)</b>
<b>Catalog Number</b>	:	33201
<b>Lot Number</b>	:	-
<b>Manufacture Date</b>	:	-
<b>Expiry Date</b>	:	3years from date of manufactured
<hr/>		
<b>Visual Attribute</b>	:	PASS
<b>Functional Test</b>	:	PASS
<b>Sterility</b>	:	We hereby certify that the goods specified above have been duly sterilized by Ethylene-Oxide gas according to ISO 11135 standard<SAL 10 <sup>-6</sup> >
<b>Cytotoxicity</b>	:	Cytotoxicity test is conducted to qualify all material resins using ISO 10993-5 standards for cytotoxicity and have been shown to be non-toxic
<b>Non-Pyrogenic</b>	:	Products are tested non-pyrogenic(LAL test) to a limit of less than 0.01EU/mL(0.4EU/device) per USP <85>
<b>RNase/DNase/DNA Free</b>	:	RNase contamination <1x10 <sup>-9</sup> Kunitz unit/uL, DNase contamination <1x10 <sup>-6</sup> Kunitz unit/uL and Human DNA <2pg
<b>BSE/TSE Free</b>	:	All products raw materials are either free of animal derived ingredients or are processed according to EMA 410/01, current revision

SPL Lifesciences Co., Ltd. hereby certifies that the product identified above have been inspected in compliance with product quality specification and requirements as documented in our ISO 9001:2015 Quality Management System (K-QA-Q031478) in Korea

Date : 01. 24. 2022

Authorized Signature : H.R. LEE



Quality Control / Assistant Manager