

EC Certificate
Directive 93/42/EEC Annex V
Production Quality Assurance
Medical Devices

Registration No.: DD 60108265 0001

Report No.: 15074616 003

Manufacturer: Wuxi Exanovo Medical
Instrument Co., Ltd.
No. 42 Xixin Road, Zhangjing
Xibei Town
Wuxi City
214194 Jiangsu
China

Products: Digital Thermometers;
Aspects of manufacture concerned with conforming of products
with the metrological requirements: Sphygmomanometers
Replaces Approval, Registration No.: DD 60097028 0001


Expiry Date: 2019-10-21

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

Effective Date: 2016-02-18

Date: 2016-02-18

Notified Body



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.