



MRI™ SureScan® pacemaker system



G70DR

Specifications

Model G70A2
Dual chamber MRI™ SureScan®
pacemaker system

vitatron • The Pace Makers

G70DR Specifications

Model G70A2

Dual chamber pacemaker system

Mechanical

Model	G70A2
Size (HxWxD mm)	44.7x47.9x7.5
M (g)	27.1
V (cc)	12.1
Connector	IS-1 BI or UNI
Radiopaque ID	V5

Battery

Type	Lithium-iodine
Voltage	2.8 V
Average projected capacity	1.3 Ah

Longevity	11.4 years*
	10.2 years†

Bradycardia Pacing

Programmable parameters

Pacing Modes	DDDR , DDD, DDIR, DDI, DVIR, DVI, DOOR, DOO, VDD, VVIR, VDIR, VVI, VDI, VVT, VOOR, VOO, AAIR, ADIR, AAI, ADI, AAT, AOO, AOO, ODO, OVO, OAO
Mode Switch	On , Off
Lower Rate	30, 35, 40... 60 ...170 min ⁻¹ (exc. 65, 85)
Upper Tracking Rate ^a	80, 90, 95... 130 ...180 min ⁻¹
Upper Sensor Rate	80, 90, 95... 130 ...180 min ⁻¹
A and RV Pulse Amplitude ^b	0.5, 0.75, 1.0... 3.5 ...4, 4.5, 5, 5.5, 6, 7.5 V
A and RV Pulse Width	0.12, 0.15, 0.21, 0.27, 0.34, 0.4 , 0.46, 0.52, 0.64, 0.76, 1, 1.25, 1.5 ms
Atrial Sensitivity	0.18, 0.25, 0.35, 0.5 , 0.7, 1, 1.4, 2, 2.8, 4 mV
Ventricular Sensitivity	1, 1.4, 2, 2.8 , 4, 5.6, 8, 11.2 mV
Pacing Polarity (A and V)	Bipolar, Unipolar, Configure
Sensing Polarity (A and V)	Bipolar, Unipolar, Configure
Paced AV (PAV)	30, 40, 50... 150 ...350 ms
Sensed AV (SAV)	30, 40, 50... 120 ...350 ms
PVARP	Auto , Varied, 150, 160, 170...500 ms
Minimum PVARP	150, 160, 170... 250 ...500 ms
PVAB	130, 140, 150... 180 ...350 ms
Atrial Refractory Period	180, 190, 200... 400 ...500 ms
Atrial Blanking Period	130, 140, 150... 180 ...350 ms
Ventricular Refractory Period	150, 160, 170... 230 ...500 ms
Ventricular Blanking (after atrial pace) (PAVB)	20, 28 , 36, 44 ms

Therapies to promote intrinsic activation

Reduced VPTM+	On , Off
Max Increase to AV	10, 20, 30... 170 ...250 ms
Sinus PreferenceTM	On , Off
Sinus Preference Zone	3, 5, 10 , 15, 20 min ⁻¹
Search Interval	5, 10, 20, 30 min
Sleep	On , Off
Sleep Rate	30, 35, 40... 50 ...90 min ⁻¹ (exc. 65, 85)
Bed Time	00:00, 00:15, 00:30... 22:00 ...23:45
Wake Time	00:00, 00:15, 00:30... 8:00 ...23:45
Single Chamber Hysteresis	Off, 40, 50, 60 min ⁻¹

Rate Response Pacing

ADL Rate	60, 65, 70... 95 ...175, 180 min ⁻¹
Rate Profile Optimization	On , Off
ADL Response	1, 2, 3 , 4, 5
Exertion Response	1, 2, 3 , 4, 5
Activity Threshold	Low, Medium Low , Medium High, High
Acceleration	15 s, 30 s , 60 s
Deceleration	2.5 min, 5 min, 10 min, Exercise
RAAV	On , Off
Start Rate	50, 55, 60... 80 ...175 min ⁻¹
Stop Rate	55, 60, 65... 120 ... 180 min ⁻¹
Maximum Offset	-10, -20, -30... -40 ...-300 ms

Rate Drop Response

Detection Type	Low Rate, Drop, Both, Off
Intervention Rate	60, 70, 75, 80... 100 ...180 min ⁻¹ (exc. 65, 85)
Intervention Duration	1, 2 , 3...15 min
Detection Beats	1, 2 , 3 beats
Drop Rate	30, 40, 50 ...100 min ⁻¹
Drop Size	10, 15, 20, 25 ...50 min ⁻¹
Detection Window	10, 15, 20, 25 , 30 s; 1, 1.5, 2, 2.5 min

Additional pacing features

PMT Intervention	On , Off
PVC Response	On , Off
Ventricular Safety Pacing	On , Off

MRI Pacing Parameters

SureScan® Pacing Mode	A00, V00, D00, ODO
SureScan Lower Rate	
Interval	60, 70, 75, 80 ... 115, 120 ^c min ⁻¹
SureScan PAV	50, 60 ... 110 ms
SureScan Atrial	
Amplitude	5.0, 5.5, 6.0, 7.5 V
SureScan Atrial	
Pulse Width	1.0, 1.25, 1.5 ms
SureScan Atrial	
Sensitivity	0.18, 0.25, 0.35, 0.5, 0.7, 1.0, 1.4, 2.0, 2.8, 4.0 mV
SureScan Ventricular	
Amplitude	5.0, 5.5, 6.0, 7.5 V
SureScan Ventricular	
Sensitivity	1.0, 1.4, 2.0, 2.8, 4.0, 5.6, 8.0, 11.2 mV
SureScan Ventricular	
Pulse Width	1.0, 1.25, 1.5 ms
SureScan Timeout Duration	24 hr
SureScan MRI Compatibility	1.5 and 3 Tesla, full body scan

Atrial Tachyarrhythmia Therapies and Interventions

Mode Switch	On , Off
Detected Rate	120, 125... 175 ...200 min ⁻¹
Detect Duration	No Delay , 10, 20...60 sec
Blanked Flutter Search	On , Off

Atrial Preference Pacing (APP) parameters

APP	On , Off
Maximum Rate (min ⁻¹)	80, 90, 95, 100 ...150
Interval Decrement (ms)	30 , 40, 50...100, 150
Search Beats	5, 10... 20 , 25, 50

Post Mode Switch Overdrive Pacing (PMOP) parameters

PMOP	On , Off
Overdrive Rate (min ⁻¹)	70, 75, 80 , 90, 95...120
Overdrive Duration (min)	0.5, 1, 2, 3, 5, 10 , 20, 30, 60, 90, 120

Nominal values indicated in **bold**

Conducted AF Response^d

Regularize V-V during AT/AF On, **Off**
Maximum Rate (min⁻¹) 80, 85, 90...**110**...130

Non-Competitive Atrial Pacing On, Off

Automatic Pacing, Sensing, and Lead Monitor

Implant Detection and Initialization

At the completion of the 30-minute Implant Detection period, Rate Profile Optimization is enabled; the appropriate pacing and sensing polarities are automatically selected by the device; Atrial and Ventricular Output Management is enabled and Amplitude and Pulse Width become adaptive. Sensing Assurance™ is enabled and Sensitivity becomes adaptive. Reduced VP™+ is enabled 60 minutes after Implant Detection is complete.

Implant Detection On/Restart, Off/Complete
Lead Monitor (A and V) Configure, Monitor Only, Adaptive (Auto Polarity Switch), Off

Notify If < **200 Ω**
Notify If > 1000, 2000, 3000, **4000 Ω**
Monitor Sensitivity 2, 3, 4 ... **8** ... 16

Atrial Output Management

Atrial Output Management Off, Monitor Only, **Adaptive**
Amplitude Margin 1.5x, **2x**, 2.5x, 3x, 4x (times)
Minimum Adapted Amplitude 0.5, 0.75...**1.5**...3.5 V
Capture Test Frequency 1, 2, 4, 8, 12 hours; Day at rest;
Day at...; 7 days at
Capture Test Time 00:00, **1:00**...23:00
Acute Phase Days Remaining Off, 7, 14, 21...84, **112**, 140, 168...
252 days

Ventricular Output Management

Ventricular Output Management Off, Monitor Only, **Adaptive**
Amplitude Margin 1.5x, **2x**, 2.5x, 3x, 4x (times)
Minimum Adapted Amplitude 0.5, 0.75...**2.0**...3.5 V
Capture Test Frequency 15, 30 min; 1, 2, 4, 8, 12 hours;
Day at rest; Day at...; 7 days at
Capture Test Time 00:00, 1:00...23:00
Acute Phase Days Remaining Off, 7, 14, 21...84, **112**, 140, 168...
252 days
V. Sensing During Search Unipolar, Bipolar, **Adaptive**

Sensing Assurance

Sensing Assurance (A and V) **On**, Off

Diagnostics

Cardiac Dashboard II

Highlights significant events, AT/AF and pacing summary, threshold and impedance trends

Atrial and ventricular pacing threshold trends
Battery longevity
Pacing summary and access to rate histogram
Atrial and ventricular lead impedance trends
Number of hours/day in atrial arrhythmia, percentage of time
Access to atrial arrhythmia diagnostics
Observations
P-wave/R-wave amplitudes and access to A and V sensitivity trends

CardioTrend™

Trend data compiles up to 6 months of daily clinical information in an easy-to-interpret graphic format

Histogram reports

Heart rate histograms
AV conduction histograms
Reduced VP™+ histogram
Sensor indicated rate profile

Atrial and ventricular episodes

Atrial and ventricular high rate episodes
Ventricular rate during atrial arrhythmias
Atrial arrhythmia durations
Multiple EGM episodes
Rate drop response episodes

Clinician selected diagnostics

Custom rate trend
Rate drop response detail
Atrial output management detail
Ventricular Output Management detail
High Rate Detail

Patient data stored in device

Patient identification
Leads implanted
Device implanted
Clinician's stored notes

Data management

Automatic printing of initial interrogation report
Full page printing
Save-to-Disk capacity for electronic file management

Follow-up and Troubleshooting

Telemetry features

Transtelephonic monitor On, Off
Extended telemetry On, Off
Extended marker Standard, Therapy Trace

Key parameter history

Initial interrogation report
Strength duration threshold test
Ventricular threshold test
Marker Channel™
Threshold margin test
Exercise test
EP studies
Magnet test
Underlying rhythm test
Sensing test
Temporary test

Magnet mode operation

	BOS	ERI
Dual chamber mode	D00 85 min ⁻¹	65
Single chamber atrial mode	A00 85 min ⁻¹	65
Single chamber ventricular mode	V00 85 min ⁻¹	65

Recommended Replacement Time (RRT) and

Elective Replacement Indicator (ERI)

Replacement message on programmer (Cardiac Dashboard II)
Battery/lead information Replacement message and battery voltage displayed on programmer
RRT and ERI initiation date Displayed on programmer



Vitatron. The Pace Makers

Vitatron - based in Europe - is the only medical device company that specializes exclusively in pacemakers. Since 1962, Vitatron pacemakers have helped restore more than 1,000,000 people in more than 60 countries to a full life. We strive to achieve perfection in everything we do. This results in unique patient-focused therapies, as well as highly cost-effective pacemakers that are easy to use.

Head Office: Vitatron Holding BV

Endepolsdomein 5,
Maastricht
NL 6229 GW
The Netherlands
www.vitatron.com

UC201708586 EE
© Vitatron Holding B.V. 2017
All Rights Reserved.

References

*DDDR or DDD 50%, 1.5 V and 2.0 V, 60 min⁻¹, 0.4 ms, 500 OHM. For Atrial Output Management the Minimum Adapted Amplitude is 1.5 V (nominal). For Ventricular Output Management, the Minimum Adapted Amplitude is 2.0 V (nominal).
†DDDR or DDD 100%, 1.5 V and 2.0 V, 60 min⁻¹, 0.4 ms, 500 OHM. For Atrial Output Management the Minimum Adapted Amplitude is 1.5 V (nominal). For Ventricular Output Management, the Minimum Adapted Amplitude is 2.0 V (nominal).

^a The atrial and ventricular Rate Limit is 200 min⁻¹ (± 20 min⁻¹).

^b Tolerance for amplitudes from 0.5 V through 6.0 V is ± 10%, and for 7.5 V is -20/+0%. Tolerances are based on 37 °C and a 500 Ω load. Amplitude is determined 200 µs after the leading edge of the pace.

^c User selection will not include 65 min⁻¹ or 85 min⁻¹.

^d Conducted AF Response is functional during Mode Switch episodes, DDIR, VVIR and VDIR modes.



G70DR • Dual chamber

vitatron • The Pace Makers