**Declaration of Conformity-V5.0** 

## Declaration of Conformity C 6 0123

Manufacturer:

Shenzhen Mindray Scientific Co., Ltd.

6/F, Bldg 2, 1203 Nanhuan Avenue, Yutang Block, Guangming District,

Shenzhen, P. R. China

**EC-Representative:** 

Shanghai International Holding Corp. GmbH (Europe)

Eiffestraße 80, 20537 Hamburg, Germany

**Product:** 

Infusion pump (Including Accessories)

Model:

BeneFusion VP5, BeneFusion VP5 Ex

Classification:

IIb (According to Rule 11 of MDD Annex IX)

**GMDN Code:** 

13215

MD Code

MD 1101

Conformity

MDD Annex II excluding (4)

**Assessment Route:** 

We herewith declare under our sole responsibility that the above mentioned products meet the provisions of the Council Directive 93/42/EEC for Medical Device, as amended by 2007/47/EC. All supporting documentations are retained under the premises of the manufacturer.

## Standards Applied:

List of (harmonized) standards for which documented evidence for compliance can be provided as attachment.

**Notified Body:** 

TÜV SÜD Product Service GmbH

Ridlerstraße 65

80339 München, Germany

Notified Body No.:

0123

Place, Date of Issue:

Shenzhen, 2019. 1.2

Signature:

Bailanhong

Name of Authorized Signatory:

Bai Yanhong

Position Held in Company:

Manager, Technical Regulation

Attachment of Declaration of Conformity: Applied Standards List-V2.0

**Product:** 

Infusion pump

Model:

BeneFusion VP5, BeneFusion VP5 Ex

**Applied Standards:** 

EN 62366-1:2015

Medical electrical equipment -- Part 2-24: Particular EN 60601-2-24:2015

requirements for the safety of infusion pumps and controllers

Medical electrical equipment - Part 1: General requirements for EN 60601-1:2006/A1:2013

basic safety and essential performance

Medical electrical equipment - Part 1-8: General requirements

for basic safety and essential performance - Collateral Standard: EN 60601-1-8:2007/A1:2013

General requirements, tests and guidance for alarm systems in

medical electrical equipment and medical electrical systems

Medical devices - Application of usability engineering to medical

devices

Medical electrical equipment - Part 1-6: General requirements

EN 60601-1-6:2010/A1:2015 for basic safety and essential performance - Collateral Standard:

Usability

Medical devices - Application of risk management to medical EN ISO 14971:2012

devices

EN 62304:2006/A1:2015 Medical device software - Software life-cycle processes

Medical Electrical Equipment - Part 1-2 General Requirements

EN 60601-1-2:2015 for Safety - Collateral Standard: Electromagnetic

compatibility-Requirements and tests

EN 1041:2008+A1:2013 Information supplied by the manufacturer of medical devices

Medical devices -

Symbols to be used with medical device labels, EN ISO 15223-1: 2016

labeling, and information to be supplied -

Part 1: General requirements