

EU Declaration of Conformity

Manufacturer: Hitachi High-Tech Corporation
Address: 1-17-1 Toranomon, Minato-ku, Tokyo 105-6409, Japan
Single Registration Number: JP-MF-000016991

European Representative: Roche Diagnostics GmbH
Address: Sandhofer Strasse 116, 68305 Mannheim, Germany
Product Name **cobas pure integrated solutions**

We, Hitachi High-Tech Corporation, declare under our sole responsibility that **cobas pure integrated solutions** (Refer to Appendix I for the components) is/are in conformity with the following European Union harmonisation legislation:

- REGULATION (EU) 2017/746 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on in vitro diagnostic medical devices and repealing Directive 98/79/EC and Commission Decision 2010/227/EU
- DIRECTIVE (EU) 2015/863 of 31 March 2015 amending Annex II to Directive 2011/65/EU of the European Parliament and of the Council as regards the list of restricted substances
- DIRECTIVE 2014/53/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 16 April 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC

Intended use/purpose: cobas pure integrated solutions is an automated analyzer, intended for running qualitative and quantitative clinical chemistry and immunochemistry assays as well as ion selective measurements.

Notified Body's name/ number (if applicable) Not applicable

IVDR conformity assessment procedures: Annex II and III of REGULATION (EU) 2017/746 (Class A)

Starting Serial No.: See Appendix II

Applied standards: See Appendix III

on behalf of the company

Date: 21-Jun-2023

DocuSigned by:


Signer Name: Yoshihiro Kawabe
Signing Reason: I approve this document
Signing Time: 21-6-2023 | 5:29:41 午後 JST
BAED48270E964289885A9C77871D4151

Yoshihiro Kawabe
General Manager
Medical Systems Quality Assurance Dept
Corporate Quality Assurance Div.
Hitachi High-Tech Corporation

Contact address:
Hitachi High-Tech Corporation
1-17-1 Toranomon, Minato-ku, Tokyo 105-6409, Japan

on behalf of the company

Date: 22-Jun-2023

DocuSigned by:

Signer Name: noda takayuki
Signing Reason: I approve this document
Signing Time: 22-Jun-2023 | 12:35:06 PM JST
32427F2C1AB9418A9A250B54ACAD1CFD

Takayuki Noda
General Manager
Medical Systems Design 1st Dept
Analytical & Medical Solution Business Group
Hitachi High-Tech Corporation

Appendix I
List of components for cobas pure integrated solutions

| Product name or component name | Basic UDI-DI | Order information | Risk classification for REGULATION (EU) 2017/746 |
|--|----------------|-------------------|--|
| sample supply unit | 761333601772B8 | 09031537001 | Class A |
| cobas e 402 analytical unit | 761333601773BA | 09031553001 | Class A |
| cobas c 303 analytical unit | 761333601771B6 | 09031529001 | Class A |
| cobas pure liquid waste container | 761333601774BC | 09033394001 | Class A |

Appendix II
List of applicable product name and serial number

REGULATION (EU) 2017/746 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on in vitro diagnostic medical devices and repealing Directive 98/79/EC and Commission Decision 2010/227/EU:

| Product name or component name | Starting serial number |
|-----------------------------------|---|
| sample supply unit | From 2201-01 onward |
| cobas e 402 analytical unit | From 2201-01 onward |
| cobas c 303 analytical unit | From 2201-01 onward (NAKA site shipment) From U301-01 onward (OMUTA site shipment) |
| cobas pure liquid waste container | Shipment from March 2022 onward |

DIRECTIVE (EU) 2015/863 of 31 March 2015 amending Annex II to Directive 2011/65/EU of the European Parliament and of the Council as regards the list of restricted substances:

| Product name or component name | Starting serial number |
|-----------------------------------|---|
| sample supply unit | From 2201-01 onward |
| cobas e 402 analytical unit | From 2201-01 onward |
| cobas c 303 analytical unit | From 2201-01 onward (NAKA site shipment) From U301-01 onward (OMUTA site shipment) |
| cobas pure liquid waste container | Shipment from March 2022 onward |

DIRECTIVE 2014/53/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 16 April 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC:

| Product name or component name | Starting serial number |
|-----------------------------------|---|
| sample supply unit | From 2201-01 onward |
| cobas e 402 analytical unit | From 2201-01 onward |
| cobas c 303 analytical unit | From 2201-01 onward (NAKA site shipment) From U301-01 onward (OMUTA site shipment) |
| cobas pure liquid waste container | Shipment from March 2022 onward |

Appendix III
List of applied standards:

REGULATION (EU) 2017/746 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 5 April 2017 on in vitro diagnostic medical devices and repealing Directive 98/79/EC and Commission Decision 2010/227/EU

| Standard number, year | Name of applied standard |
|--|---|
| EN ISO 13485: 2016 | Medical devices – Quality management systems - Requirements for regulatory purposes |
| EN ISO 14971: 2019 | Medical devices - Application of risk management to medical devices |
| IEC 62366-1:2015 +AMD 1:2020 | Medical devices - Part 1: Application of usability engineering to medical devices - Amendment 1 |
| EN 13612: 2002 | Performance evaluation of in vitro diagnostic medical devices |
| EN ISO 18113-1: 2011 | In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 1: Terms, definitions and general requirements |
| EN ISO 18113-3: 2011 | In vitro diagnostic medical devices - Information supplied by the manufacturer (labeling) - Part 3: In vitro diagnostic instruments for professional use |
| EN ISO 15223-1:2021 | Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements |
| IEC 61010-2-101: 2015 | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment |
| IEC 61326-2-6: 2012/ EN 61326-2-6:2013 | Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 2-6: Particular requirements – In vitro diagnostic (IVD) medical equipment |

DIRECTIVE (EU) 2015/863 of 31 March 2015 amending Annex II to Directive 2011/65/EU of the European Parliament and of the Council as regards the list of restricted substances:

| Standard number, year | Name of applied standard |
|-----------------------|--|
| EN IEC 63000:2018 | Technical documentation for the assessment of electrical and electronic products with respect to the restriction of hazardous substances |

DIRECTIVE 2014/53/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 16 April 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC:

| Standard number, year | Name of applied standard |
|-----------------------|--|
| EN 300 330 V2.1.1 | Short Range Devices (SRD); Radio equipment in the frequency range 9 kHz to 25 MHz and inductive loop systems in the frequency range 9 kHz to 30 MHz |
| IEC61010-2-101:2002 | Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-101: Particular requirements for in vitro diagnostic (IVD) medical equipment |
| EN 62479:2010: | Assessment of the compliance of low power electronic and electrical equipment with the basic restrictions related to human exposure to electromagnetic fields (10 MHz to 300 GHz) |
| EN 301 489-1 V1.9.2: | Electro Magnetic Compatibility (EMC) standard for radio equipment and services; Part 1: Common technical requirements |
| EN 301 489-3 V1.6.1: | Electromagnetic compatibility and Radio spectrum Matters (ERM); Electro-Magnetic Compatibility (EMC) standard for radio equipment and services; Part 3: Specific conditions for Short-Range Devices (SRD) operating on frequencies between 9 kHz and 246 GHz |

End of the document