



Centrum für Hygiene
und medizinische Produktsicherheit

Certificate

for the company

Getinge Disinfection AB
Ljungadalsgatan 11
P O Box 1505
S- 35115 Växjö

manufactured the

Washer Disinfector
“WD 15 CLARO”

can be certified that this machine fulfills the requirements and

therefore the type test of

EN ISO 15883-1 (2009)

“General requirements, definitions and tests”.

and

EN ISO 15883-2 (2009)

“Requirements and tests for washer-disinfectors employing thermal disinfection for surgical instruments, anaesthetic equipment, bowls, dishes, receivers, utensiles, glassware, etc.”

Tested in program “Surgical instruments”

Identification Number: 99234

Schwerin, 02.05.2011

Dipl.-Ing (FH) Margrit Werner
Executive Manager



Prof. Dr. med. H.-P. Werner
Scientific-Technical Director