医疗器械质量管理体系认证证书

证书号: HXC18921YL008R0M

兹 证 明 山东博科生物产业有限公司

统一社会信用代码: 9137018179889855X7

注册地址: 山东省济南市章丘区明水经济开发区(经十东路与明埠路交界山东博科产业园) 办公地址: 山东省济南市章丘区明水经济开发区(经十东路与明埠路交界山东博科产业园) 生产/经营地址: 山东省济南市工业南路 51 号(生产)/济南市章丘明水经济开发区经十东路与明埠路交界

山东博科产业园(研发、生产、销售)/济南市章丘明水经济开发区清河大街 2 号鑫岳工业园 8# (生产)

/山东省济南市平阴县锦水高端装备制造产业园 3 号标准厂房(生产)

邮编: 250200

医疗器械质量管理体系符合

ISO 13485: 2016 标准要求

通过认证的范围为

用于体外诊断试剂盒、生物安全柜、医用气溶胶吸附器、血液冷藏箱、医用恒温箱、全自动酶免工作站、血细胞分析用稀释液、血细胞分析用溶血剂、洗板机、医用冷藏箱、全自动生化分析仪、全自动化学发光测定仪、核酸提取仪、全自动核酸提取仪、一次性使用病毒采样管、医用离心机的研发、生产和销售(资质许可范围内

第一次监审

贴标处

第二次监审

贴标处

第三次监审

贴标处

证书有效期内,获证组织须每年至少接受一次监督审核并加贴合格标志,证书方为有效。本证书信息可在国家认证认可监督管理委员会官方网站:www.cnca.gov.cn查询。证书有效性以右侧二维码扫描内容为准





HEND.



有效日期: 2024年7月4日 010-57146599

MEDICAL DEVICES QUALITY MANAGEMENT SYSTEMS CERTIFICATE

Certification No.: HXC18921YL008R0M

Awarded to Biobase Biodustry(Shandong)Co., Ltd.

Organization Code: 9137018179889855X7

Registration Add.: Mingshui Economic Development Zone, Zhangqiu District, Jinan City, Shandong Province (Biobase Industrial Park in a intersection of East Jingshi Road and Mingbu Road)

Office Add.: Mingshui Economic Development Zone, Zhangqiu District, Jinan City, Shandong Province (Biobase Industrial Park in a intersection of East Jingshi Road and Mingbu Road)

Production/Service Add.: No.51 Industrial South road, Jinan City, Mingshui Economic Development Zone, Zhangqiu District, Jinan City(Biobase Industrial Park in a intersection of East Jingshi Road and Mingbu Road), No.2 Qinghe street, Zhangqiu District, Jinan City, Standard Plant No. 3, Jinshui High-end Equipment Manufacturing Industrial Park, Pingyin County, Jinan City, Shandong Province

Zip Code: 250200

The Medical Devices Quality Management Systems of the above Organization has been assessed and Found to be in accordance with the requirements of the standard:

ISO 13485:2016

Scope of this management Certification

R&D, production and sales of In Vitro Diagnostic Reagents, Biological Safety Cabinets, Air Purifier, Blood Bank Refrigerator, Medical Incubator, Automatic ELISA Processor, Hematology Analyzer Diluent, Hematology Analyzer Lyse, Microplate Washer, Laboratory Refrigerator, Automatic chemistry analyzer, Automatic Chemiluminescence Analyzer, Automatic Nucleic acid extration system, Disposable Virus Sampling Tube, Medical centrifuge (within the scope of qualification license)

First S.A. Labelling

Second S.A. Labelling

Third S.A. Labelling

The certified organization must accept at least one surveillance audit annually and a qualified mark shall be attached onto the certificate, within validity period of the certificate, then the certificate can be maintained valid.

For more details of this certificate, please login CNCA's website: www.cnca.gov.cn.

For the validity of the certificate, please scan the QR codes on the right ride for accurate information.







