

ACON BIOTECH (HANGZHOU) CO.,LTD
CERTIFICATE OF ANALYSIS

PRODUCT NAME: Urinalysis Reagent Strips

PURCHASE NO.: 37740066C

BATCH NO.: URS0030251

EXPIRATION DATE: 2022-03-28

QUANTITY: 100x100T

ANALYSIS		SPECIFICATIONS (Color card)	TEST RESULT
Glucose	Neg. urine	Negative	Complies
	250mg/dL	>100mg/dL <500mg/dL	Complies
Bilirubin	Neg. urine	Negative	Complies
	Yihua Urine controls	>++	Complies
Ketone	Neg. urine	Negative	Complies
	Yihua Urine controls	>40mg/dL <160mg/dL	Complies
S.G	1.000	$\geq 1.000 < 1.005$	Complies
	1.025	$> 1.020 < 1.030$	Complies
Blood	Neg. urine	Negative	Complies
	Yihua Urine controls	$\geq ++ \leq +++$	Complies
pH	5.0	$\geq 5.0 < 6.0$	Complies
	9.0	$> 8.0 \leq 9.0$	Complies
Protein	Neg. urine	Negative	Complies
	Yihua Urine controls	>+, $\leq +++$	Complies
Urobilinogen	Neg. urine	Negative	Complies
	4~12 mg/dL	>4mg/dl $\leq 12\text{mg/dl}$	Complies
Nitrite	Neg. urine	Negative	Complies
	Yihua Urine controls	+	Complies
Leukocyte	Neg. urine	Negative	Complies
	Yihua Urine controls	>+, <+++	Complies

Quality Department: Among. Hu

Date: July. 02, 2020