

Handrop® digital - SC

Single channel variable volume digital pipettes with digital display.
Precise volume selector.
Easy calibración and maintenance.
Lightweight and ergonomic design with tip ejector.

| code | volume range µl | increments µl | accuracy % | precision % | case weight | case volume |
|--------|-----------------|---------------|------------|-------------|-------------|-------------|
| 770005 | 0.1 - 2.5 | 0.05 | 2.5 - 12 | 2 - 5 | 0.19 | 0.001 |
| 770000 | 0.5 - 10 | 0.1 | 1.0 - 2.5 | 0.8 - 1.5 | 0.19 | 0.001 |
| 770080 | 2 - 20 | 0.5 | 0.9 - 3.0 | 0.4 - 2.0 | 0.19 | 0.001 |
| 770050 | 10 - 100 | 1 | 0.8 - 3.0 | 0.15 - 1.5 | 0.20 | 0.001 |
| 770070 | 20 - 200 | 1 | 0.6 - 3.0 | 0.15 - 1.0 | 0.19 | 0.001 |
| 770060 | 100 - 1,000 | 5 | 0.6 - 2.0 | 0.2 - 0.7 | 0.20 | 0.001 |
| 770040 | 200 - 1,000 | 5 | 0.6 - 0.9 | 0.2 - 0.3 | 0.19 | 0.001 |
| 770110 | 1 - 5 ml | 0.05 ml | 0.5 - 0.7 | 0.15 - 0.3 | 0.21 | 0.001 |



Handrop® digital - MC

8 and 12 channel variable volume digital pipettes with digital display.
Precise volume selector.
Easy calibración and maintenance.
Lightweight and ergonomic design with tip ejector.

| code | volume range µl | increments µl | accuracy % | precision % | case weight | case volume |
|--------------------|-----------------|---------------|------------|-------------|-------------|-------------|
| 8 channels | | | | | | |
| 770210 | 0.5 - 10 | 0.1 | 1.0 - 4.0 | 1.5 - 4.0 | 0.36 | 0.0015 |
| 770220 | 5 - 50 | 0.5 | 1.0 - 3.0 | 0.5 - 2.0 | 0.36 | 0.0015 |
| 770240 | 50 - 300 | 5 | 0.7 - 1.5 | 0.25 - 0.8 | 0.36 | 0.0015 |
| 12 channels | | | | | | |
| 770310 | 0.5 - 10 | 0.1 | 1.0 - 4.0 | 1.5 - 4.0 | 0.30 | 0.0015 |
| 770320 | 5 - 50 | 0.5 | 1.0 - 3.0 | 0.5 - 2.0 | 0.38 | 0.0015 |
| 770340 | 50 - 300 | 5 | 0.7 - 1.5 | 0.25 - 0.8 | 0.36 | 0.0015 |



Handrop® stepper

Single channel stepper.
Allows to dispense up to 40 different volumes and up to 48 doses.
Compatible with tips on the page 192.
Volumes from 1 µl to 5 ml can be dispensed.

| code | description | case weight | case volume |
|--------|-----------------|-------------|-------------|
| 721008 | Handrop Stepper | 0.20 | 0.0015 |



QUALITY CERTIFICATE

CUSTOMER ...: 2988 - GBG MOLDOVA

INVOICE: FV00204411

Deltalab hereby certifies that the below mentioned product/s has/have been manufactured and/or commercialized under the requirements of the Standard ISO 9001:2015 "Quality management systems".

The medical devices have also been manufactured and/or commercialized under the requirements of the standard ISO 13485:2016 "Medical devices. Quality management systems. Requirements for regulatory purposes".

| Code | Description | Batch | Expire |
|---------|---------------------------------|----------------|--------|
| 770050P | MICROPIPETTE PLUS- SC 10-100ul | YM247BF0004134 | |
| 770240 | DIGITAL HANDROP MC8 50-300UL | GJ238AD0000161 | |
| 770110P | MICROPIPETTE PLUS - SC 1-5ML | YM24CBJ0024037 | |
| 770110P | MICROPIPETTE PLUS - SC 1-5ML | YM24CBJ0024027 | |
| 770110P | MICROPIPETTE PLUS - SC 1-5ML | YM24CBJ0024020 | |
| 770110P | MICROPIPETTE PLUS - SC 1-5ML | YM24CBJ0024005 | |
| 400700 | TUBE 13X75 PP | 32501173 | 2/30 |
| 200410 | EPPENDORF TYPE MICROTUBE ORANGE | 3724 | |
| 200405 | EPPENDORF TYPE MICROTUBE BLUE | 0822 | |
| 200401 | EPPENDORF TYPE MICROTUBE YELLOW | 1723 | |
| M-021V | COMBI RACK GREEN | 24441 | |

Date: 16/4/2025

Quality Department

Certificate issued automatically. Valid without signature.

This is a translation of the certificate ES16/20725.01



DELTALAB, S.L.

Pol.Ind. La Llana, Plaza de la Verneda, 1, 08191 Rubí, Barcelona

Has been assessed under the management system of the certified organisation defined in the main certificate ES16/20725 as meeting the requirements of

ISO 9001:2015

For the following activities

Design, manufacture and sale of laboratory material for the collection, transport and conservation of samples for microbiological, molecular biology, hematology, biochemistry, histology, microscopy and colorimetric analysis.

Commercialization of equipment for the storage of prepared samples, cryogenic stored samples, general labware and industrial packages.

Commercialization of equipment and instrumentation for the laboratory, diagnostic kits, healthcare products and cosmetics.

This certificate is valid from 11 October 2022 until 11 October 2025 and remains valid subject to satisfactory surveillance audits.

Issue 2.

The validity of this certificate depends on the validity of the main certificate.

SGS International Certification Services Iberica, S.A.U.

C/Trespaderne, 29. 28042 Madrid. España

t +34 91 313 8115 - www.sgs.com



This document is an authentic electronic certificate for Client' business purposes use only. Printed version of the electronic certificate are permitted and will be considered as a copy. This document is issued by the Company subject to SGS General Conditions of certification services available on Terms and Conditions | SGS. Attention is drawn to the limitation of liability, indemnification and jurisdictional clauses contained therein. This document is copyright protected and any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful.



This is a translation of the certificate ES19/86440.01

SGS

DELTALAB, S.L.

Pol.Ind. La Llana, Plaza de la Verneda, 1, 08191 Rubí, Barcelona

Has been assessed under the management system of the certified organisation defined in the main certificate ES19/86440 as meeting the requirements of

ISO 14001:2015

For the following activities

Design, manufacture and sale of laboratory material for the collection, transport and conservation of samples for microbiological, molecular biology, hematology, biochemistry, histology, microscopy and colorimetric analysis.

Commercialization of equipment for the storage of prepared samples, cryogenic stored samples, general labware and industrial packages.

Commercialization of equipment and instrumentation for the laboratory, diagnostic kits, healthcare products and cosmetics.

This certificate is valid from 31 August 2022 until 29 August 2025 and remains valid subject to satisfactory surveillance audits.

Issue 2.

The validity of this certificate depends on the validity of the main certificate.

SGS International Certification Services Iberica, S.A.U.

C/Trespaderne, 29. 28042 Madrid. España

t +34 91 313 8115 - www.sgs.com



This document is an authentic electronic certificate for Client' business purposes use only. Printed version of the electronic certificate are permitted and will be considered as a copy. This document is issued by the Company subject to SGS General Conditions of certification services available on [Terms and Conditions](#) | [SGS](#). Attention is drawn to the limitation of liability, indemnification and jurisdictional clauses contained therein. This document is copyright protected and any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful.

