

Catalog Number T10282
Product Name Trypan Blue stain 0.4% *for use with Countess™ automated cell counter*
Appearance blue solution
Lot Number 3396759A

Customer Use Suggestion: Filtration, centrifugation, or warming to 37C for 15 minutes can be useful to minimize naturally occurring precipitates.

	LOT DATA	SPECIFICATION
ASSAY Osmolality	307 mOsm/kg	≥290 to ≤340 mOsm/kg
ASSAY Sterility Testing	meets specification	negative
MICROSCOPY Viable Stain	meets specification	acceptable



Zach Luedtke, Quality Assurance Manager
26-Aug-2025

Life Technologies Corporation certifies on the date above that this is an accurate record of the analysis of the subject lot, and that the data conform to the specifications in effect for this product at the time of analysis.

Products are under warranty for one year from the date of shipment unless otherwise stated on this document or the product literature.