

CERTIFICATE OF ANALYSIS

Certificado de Análisis / Certificat de contrôle / Chargen-Prüfprotokoll
RealPCR ASFV DNA Mix [ASFV DNA Mix]

Product and lot number: 99-56020 - 240188

Expiration date: 22-FEB-2026

Number of reactions: 100

Manufacturing date: AUG-2024

The RealPCR ASFV DNA Mix needs to be used with the insert 06-56020-04.

The PC compatibility is ≥ 1.3 and IPC compatibility is ≥ 1.2 .

Each RealPCR ASFV DNA Mix is tested with RealPCR DNA Master Mix on release panel (2 replicates of each panel member) and must meet the specifications detailed below.

Functional Test	Specification (mean Ct)	Result (Pass/ Fail)
ASFV 100 copy / reaction	< 36	Pass
ASFV 1000 copy / reaction	< 32	Pass
Internal Control	< 33	Pass
No template control	no Ct value	Pass

This product was performance tested and has met all quality control specifications required for release.
This information is released by:

Name: Rachel Mauran, Quality Specialist

Signature:

