

Anexa tehnica **Lot 1. Distilator cu mono distilare 12L**

**ND 12, Nuve, Turcia**

	<b>Parametri ceruti</b>	<b>Parametri oferiti</b>
Descriere	Aparat electric ce produce apa distilata de inalta calitate, chiar si in conditiile utilizarii unei surse de apa de calitate scazuta cu domeniul de utilizare în laborator, farmacie, etc.	Aparat electric ce produce apa distilata de inalta calitate, chiar si in conditiile utilizarii unei surse de apa de calitate scazuta cu domeniul de utilizare în laborator, farmacie, etc.
Parametru	Specificație	
Capacitatea de distilare	≥ 12 l/h	12 l/h
Materialul carcasei	otel inox/otel vopsit/aluminiu vopsit	otel inox/otel vopsit/aluminiu vopsit
Materialul de contact si elemente de incalzire	oțel inox	oțel inox
Conductivitate	≤ 2.5 μS/cm	2.5 μS/cm
Modul de operare	continuu	continuu
Protectie in cazul intreruperii accidentale a alimentarii cu apa	obligatoriu	obligatoriu
Alimentare	380 V- 50 Hz	380 V- 50 Hz
Sistem de filtrare a apei cu sistem de fixare pe perete, inclus cu toate componentele necesare pentru montare		Filtru DP 10 trio 3/4"
Filtru pentru filtrare mecanică, 1μm	min. 8 buc.	K1 cartus de schimb (1mcr)
Filtru anticalcar cu polifosfați (cu granule)	min. 2 rezervi de granule	2x Poliphosfat Media ( ambalaj 0,5 kg)
Filtru de clor	min. 4 buc.	4x K5 AN cartus de schimb (5 mcr)
Incalzitoare de rezerva	da, min 1 set	1 set
Garanție	≥ 36 luni	36 luni
Livrarea, instalarea și testarea la locul necesar fără implicarea beneficiarului	da	Da
Asigurarea tuturor elementelor necesare pentru instalarea completă, funcțională și sigură a sistemului, inclusiv cele aferente conectivității, integrării în infrastructura existentă.	da	da
Preluarea ambalajelor (cutiile) în care a fost livrat sistemul/dispozitivul	da, la solicitarea beneficiarului	da

Ghid rapid	Min. 1 buc, ghid rapid al utilizatorului în limba de stat (Română), laminată și atașată de dispozitiv. Disponibil atât în format fizic (pe hârtie), cât și în format PDF.	1 buc, ghid rapid al utilizatorului în limba de stat (Română), laminată și atașată de dispozitiv. Disponibil atât în format fizic (pe hârtie), cât și în format PDF.
Manuale de utilizare	Disponibil atât în format fizic (pe hârtie), cât și în format PDF. În limba de stat (română). În cazul în care nu este disponibil în limba de stat, trebuie prezentată o traducere oficială a manualului, certificată prin ștampila unui birou de traduceri autorizat.	Disponibil atât în format fizic (pe hârtie), cât și în format PDF. În limba de stat (română). În cazul în care nu este disponibil în limba de stat, trebuie prezentată o traducere oficială a manualului, certificată prin ștampila unui birou de traduceri autorizat.
Manual de service	Disponibil atât în format fizic (pe hârtie), cât și în format PDF. În limba de stat (Română) sau în una din limbile de circulație internațională (rusă sau engleză) conform Legii nr. 102 din 09.06.2017, cap. 4, art. 14, al. (3) și art. 15, al. (6).	Disponibil atât în format fizic (pe hârtie), cât și în format PDF. În limba de stat (Română) sau în una din limbile de circulație internațională (rusă sau engleză) conform Legii nr. 102 din 09.06.2017, cap. 4, art. 14, al. (3) și art. 15, al. (6).
Training pentru utilizatori	Da, pregătire de lucru, mod de utilizare a dispozitivului, întreținerea zilnică.	Da, pregătire de lucru, mod de utilizare a dispozitivului, întreținerea zilnică.
Training pentru personal tehnic	Ajustarea, calibrarea, remedierea defecțiunilor minore	Ajustarea, calibrarea, remedierea defecțiunilor minore



# ND SERIES

## Water Distillers



- New design with distilled water storage tank
- Fully automatic operation with microprocessor control system
- Exhaust port for CO<sub>2</sub> gas disposal
- Safety features with warning lights: - Water level is high - Water is insufficient - Distillation - Heater malfunction (ND 12) - Distilled tank full - Distilled tank half
- Automatic control system that stops the operation of the device when the distilled water tank is full and starts it again when the water level in the tank is halved
- Energy savings by reusing a certain amount of heated water
- Heaters protected against waterless operation
- Long-lasting stainless steel heaters
- Stainless steel boiling pot, storage tank and condensation coil
- Design that allows both tabletop and wall-mounted use
- Siliphos filter that minimizes calcification of heaters

### TECHNICAL SPECIFICATIONS

	ND 4	ND 12
Distilled water capacity lt./hr.	4	12
Storage tank capacity liters	8	24
Cooling water consumption lt./hr.	40	120
Power consumption	3x1000 W	3x3000 W
Power supply	230 V - 50 Hz	400 V - 50 Hz, 3 faz+N+G
Internal Material	Stainless Steel	
External Material	Epoxy polyester powder coated aluminium	
External Dimensions ( WxDxH) mm	820x395x650	820x395x710
Packing Dimensions ( WxDxH) mm	880x480x810	880x480x860
Net / Packed Weight kg.	40 / 52	45 / 57



**NUVE SANAYİ MALZEMELERİ İMALAT VE TİCARET A.Ş.**

Saracalar Mah. Saracalar Kümeevleri No: 4/2 Akyurt 06750 ANKARA / TURKEY t. +90 312 399 28 30 f. +90 312 399 21 97

[nuve.com.tr](http://nuve.com.tr) [sales@nuve.com.tr](mailto:sales@nuve.com.tr)

ISO 9001: 2015

ISO 13485: 2016

ISO 45001: 2018





**Date: 28.07.2025**

**Outgoing No: NV2025200**

To whom it may concern,

We, Nüve Sanayi Malzemeleri İmalat ve Ticaret A.Ş., the manufacturer of the ND series water distillers and other laboratory equipment, confirm that the conductivity of the distilled water produced by the ND12 model is approximately 2.5  $\mu$ S/cm at 20 °C.

Please note that the conductivity may vary depending on the quality of the inlet water.

Best regards,  
Nüve Sanayi Malzemeleri İmalat ve Ticaret A.Ş.

**NÜVE**  
SANAYİ MALZEMELERİ İMALAT VE TİCARET A.Ş.  
Saracalar Mahallesi Saracalar Kümeevleri No: 4/2  
06750 Akyurt/ANKARA / TÜRKİYE  
Kavaklıdere V.C. 0632 005 0327  
Mersis No: 0632 0050 3270 0019

#### **NÜVE SANAYİ MALZEMELERİ İMALAT VE TİCARET A.Ş.**

##### **Head Office& Factory:**

Saracalar Mah. Saracalar Kümeevleri  
No: 4/2 Akyurt 06750 ANKARA / TURKEY  
t. +90 312 399 28 30 (pbx)  
f. +90 312 399 21 97

##### **İstanbul Showroom:**

Turgut Özal Cad. Karagül İş Merkezi  
No: 84/89 Fındıkzade 34280 İSTANBUL / TURKEY  
t. +90 212 532 07 47  
f. +90 212 531 23 28

##### **İzmir Showroom::**

Mansuroğlu Mah. Ankara Cad.  
179/A Bayraklı 35030 İZMİR / TURKEY  
t. +90 232 445 25 70  
f. +90 232 445 57 80

**Ticaret Sicil No:**  
Ankara/18221

**Mersis No:**  
0632-0050-3270-0019

# EC DECLARATION OF CONFORMITY

We, NÜVE Sanayi Malzemeleri İmalat ve Ticaret A.Ş.  
Saracalar Mah. Saracalar Kümeevleri No: 4/2 (Eski Esenboğa Yolu 22 Km.)  
Akyurt 06750 Ankara / TURKEY

Herewith declare that the following described product meets the provisions of the following directives. All supporting documentation is retained under the premises of the manufacturer.

PRODUCT:

WATER DISTILLER

MODEL:

ND 4 – ND 8 – ND 12

APPLICABLE HARMONISED  
STANDARDS:

EN 61010-1: 2010  
EN 61000-6-3: 2007  
EN 61326-1: 2013

SIGNATURE:

  
Beril İZGIN  
General Manager

CE



CERTIFICATE

# NÜVE SANAYİ MALZEMELERİ İMALAT VE TİCARET ANONİM ŞİRKETİ

Saracalar Mahallesi Saracalar Kümeevleri No: 4/2 Akyurt Ankara, Türkiye

**Design, Manufacture, Sales And After Sales Services Of Laboratory  
And Sterilization Equipments**

with a scope of

## ISO 9001:2015

Has established a quality management system in accordance  
with international standard.

*"Following elements of the standard are excluded"*  
*"None"*

Certificate No : M 8215  
Initial Certification Date : 28 April 2010  
Certification Date : 30 January 2025  
Expiration Date : 29 January 2028

Kiwa Belgelendirme Hizmetleri A.Ş.  
ITOSB 9. Cadde No. 15 Tepeören Tuzla  
Istanbul / Turkey

Tel: + 90 216 593 25 75  
Faks: + 90 216 593 25 74  
[info@kiwa.com.tr](mailto:info@kiwa.com.tr)  
[www.kiwa.com.tr](http://www.kiwa.com.tr)

Certificate is valid till expiration date,  
subject to successful completion of  
periodical surveillance audits.  
Please contact above numbers for  
detailed information.

General Manager



TÜRKAK BDS NO  
YS-A194-43FA

# NÜVE SANAYİ MALZEMELERİ İMALAT VE TİCARET ANONİM ŞİRKETİ

Saracalar Mahallesi Saracalar Kümeevleri No: 4/2 Akyurt Ankara, Türkiye

**Design, Manufacture, Sales And After Sales Services Of Laboratory  
And Sterilization Equipments**

with a scope of

## EN ISO 13485:2016

Has established a management system in accordance  
with international Medical Devices Quality Management System Standard

*"Following elements of the standard are excluded"*

*"7.5.5" "7.5.7" "7.5.9.2"*

Certificate No	: M 8216
Initial Certification Date	: 28 April 2010
Certification Date	: 30 January 2025
Expiration Date	: 29 January 2028

Kiwa Belgelendirme Hizmetleri A.Ş.  
İTOSB 9. Cadde No. 15 Tepeören Tuzla  
İstanbul / Turkey

Tel: +90 216 593 25 75  
Faks: +90 216 593 25 74  
[info@kiwa.com.tr](mailto:info@kiwa.com.tr)  
[www.kiwa.com.tr](http://www.kiwa.com.tr)

Certificate is valid till expiration date,  
subject to successful completion of  
periodical surveillance audits.  
Please contact above numbers for  
detailed information.



General Manager



TÜRKAK BDS NO  
YS-843F-FB1A